

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
 Data File : BN011000.D
 Acq On : 20 Jun 2020 11:35
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 22 00:31:18 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 19 12:43:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	0.00
2	1,4-Dioxane	0.400	0.498	-24.5	146	0.00
3	n-Nitrosodimethylamine	0.400	0.489	-22.2	171	0.02
4 S	2-Fluorophenol	0.400	0.461	-15.3	139	0.00
5 S	Phenol-d6	0.400	0.450	-12.5	138	0.00
6	bis(2-Chloroethyl)ether	0.400	0.470	-17.5	144	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	129	0.00
8 S	Nitrobenzene-d5	0.400	0.436	-9.0	135	0.00
9	Naphthalene	0.400	0.420	-5.0	129	0.00
10	Hexachlorobutadiene	0.400	0.431	-7.7	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.418	-4.5	128	0.00
12	2-Methylnaphthalene	0.400	0.422	-5.5	132	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	135	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.403	-0.8	141	0.00
15 S	2-Fluorobiphenyl	0.400	0.420	-5.0	133	0.00
16	Acenaphthylene	0.400	0.416	-4.0	139	0.00
17	Acenaphthene	0.400	0.411	-2.7	134	0.00
18	Fluorene	0.400	0.414	-3.5	136	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	154	0.00
20	4-Bromophenyl-phenylether	0.400	0.394	1.5	148	0.00
21	Hexachlorobenzene	0.400	0.400	0.0	147	0.00
22	Pentachlorophenol	0.400	0.360	10.0	143	0.00
23	Phenanthrene	0.400	0.394	1.5	148	0.00
24	Anthracene	0.400	0.366	8.5	142	0.00
25 SURR	Fluoranthene-d10	0.400	0.346	13.5	132	0.00
26	Fluoranthene	0.400	0.346	13.5	133	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	144	-0.01
28	Pvrene	0.400	0.386	3.5	134	0.00
29 S	Terphenyl-d14	0.400	0.378	5.5	131	0.00
30	Benzo(a)anthracene	0.400	0.409	-2.2	149	0.00
31	Chrysene	0.400	0.398	0.5	133	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.436	-9.0	161	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	131	0.00
34 I	Perylene-d12	0.400	0.400	0.0	136	0.00
35	Benzo(b)fluoranthene	0.400	0.385	3.8	126	0.00
36	Benzo(k)fluoranthene	0.400	0.389	2.8	126	0.00
37 C	Benzo(a)pyrene	0.400	0.405	-1.3	137	0.00
38	Dibenzo(a,h)anthracene	0.400	0.396	1.0	132	0.00
39	Benzo(a,h,i)perylene	0.400	0.394	1.5	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN061920\
Data File : BN011000.D
Acq On : 20 Jun 2020 11:35
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 22 00:31:18 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 19 12:43:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				