


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061920\
 Data File : BN010982.D
 Acq On : 20 Jun 2020 00:44
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 19 12:43:49 2020



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2474

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.0	28893	PASS
68	69	0.00	2	1.5	759	PASS
69	198	0.00	100	40.7	49045	PASS
70	69	0.00	2	0.4	210	PASS
127	198	10	80	50.1	60424	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	120608	PASS
199	198	5	9	6.9	8354	PASS
275	198	10	60	27.8	33568	PASS
365	198	1	100	4.3	5146	PASS
441	198	0.01	100	10.9	13123	PASS
442	198	50	100	72.0	86829	PASS
443	442	15	24	17.1	14859	PASS