

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061924\
 Data File : BN032069.D
 Acq On : 19 Jun 2024 14:29
 Operator : MA/JU
 Sample : P2928-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-196-GW-540-542

Quant Time: Jun 19 15:13:24 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 11:30:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.646	152	7152	0.400	ng	0.00
7) Naphthalene-d8	10.421	136	24274	0.400	ng	0.00
13) Acenaphthene-d10	14.285	164	15743	0.400	ng	0.00
19) Phenanthrene-d10	17.029	188	34544	0.400	ng	# 0.00
29) Chrysene-d12	21.238	240	27996	0.400	ng	0.00
35) Perylene-d12	23.458	264	28932	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.248	112	2349	0.116	ng	0.00
5) Phenol-d6	6.830	99	2202	0.083	ng	0.00
8) Nitrobenzene-d5	8.798	82	4491	0.223	ng	0.00
11) 2-Methylnaphthalene-d10	12.017	152	9636	0.249	ng	0.00
14) 2,4,6-Tribromophenol	15.787	330	1509	0.200	ng	0.00
15) 2-Fluorobiphenyl	12.906	172	19058	0.319	ng	0.00
27) Fluoranthene-d10	19.072	212	19402	0.206	ng	0.00
31) Terphenyl-d14	19.676	244	19244	0.281	ng	0.00
Target Compounds						
25) Phenanthrene	17.078	178	2546	0.027	ng	Qvalue 96
28) Fluoranthene	19.100	202	2352	0.020	ng	# 80
30) Pyrene	19.467	202	2275	0.020	ng	# 94
34) Bis(2-ethylhexyl)phtha...	21.148	149	13935	0.194	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061924\
Data File : BN032069.D
Acq On : 19 Jun 2024 14:29
Operator : MA/JU
Sample : P2928-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-196-GW-540-542

Quant Time: Jun 19 15:13:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 11:30:34 2024
Response via : Initial Calibration

