

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061924\
 Data File : BN032074.D
 Acq On : 19 Jun 2024 18:50
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 19 19:17:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 11:30:34 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	66	0.00
2	1,4-Dioxane	0.400	0.346	13.5	66	0.00
3	n-Nitrosodimethylamine	0.400	0.341	14.8	66	0.00
4 S	2-Fluorophenol	0.400	0.332	17.0	63	0.00
5 S	Phenol-d6	0.400	0.338	15.5	65	0.00
6	bis(2-Chloroethyl)ether	0.400	0.339	15.3	65	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	71	0.00
8 S	Nitrobenzene-d5	0.400	0.334	16.5	69	0.00
9	Naphthalene	0.400	0.380	5.0	77	0.00
10	Hexachlorobutadiene	0.400	0.397	0.8	80	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.396	1.0	81	0.00
12	2-Methylnaphthalene	0.400	0.398	0.5	81	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	75	0.00
14 S	2,4,6-Tribromophenol	0.400	0.308	23.0	72	0.00
15 S	2-Fluorobiphenyl	0.400	0.380	5.0	83	0.00
16	Acenaphthylene	0.400	0.342	14.5	77	0.00
17	Acenaphthene	0.400	0.374	6.5	82	0.00
18	Fluorene	0.400	0.376	6.0	82	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	71	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.311	22.3	48	0.00
21	4-Bromophenyl-phenylether	0.400	0.389	2.8	81	0.00
22	Hexachlorobenzene	0.400	0.414	-3.5	85	0.00
23	Atrazine	0.400	0.308	23.0	65	0.00
24	Pentachlorophenol	0.400	0.364	9.0	62	0.00
25	Phenanthrene	0.400	0.391	2.3	81	0.00
26	Anthracene	0.400	0.359	10.3	77	0.00
27 SURR	Fluoranthene-d10	0.400	0.366	8.5	76	0.00
28	Fluoranthene	0.400	0.370	7.5	77	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	60	0.00
30	Pyrene	0.400	0.412	-3.0	74	0.00
31 S	Terphenyl-d14	0.400	0.423	-5.7	77	0.00
32	Benzo(a)anthracene	0.400	0.368	8.0	67	0.00
33	Chrysene	0.400	0.384	4.0	69	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.324	19.0	58	0.00
35 I	Perylene-d12	0.400	0.400	0.0	48	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	60	0.00
37	Benzo(b)fluoranthene	0.400	0.422	-5.5	60	0.00
38	Benzo(k)fluoranthene	0.400	0.430	-7.5	62	0.00
39 C	Benzo(a)pyrene	0.400	0.397	0.8	58	0.00
40	Dibenzo(a,h)anthracene	0.400	0.405	-1.3	61	0.00
41	Benzo(g,h,i)perylene	0.400	0.401	-0.3	61	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4