

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN061925\
 Data File : BN037320.D
 Acq On : 18 Jun 2025 18:19
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN061925

Quant Time: Jun 19 02:39:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN061925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 18 18:40:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	151	0.00
2	1,4-Dioxane	0.400	0.501	-25.2#	168	0.00
3	n-Nitrosodimethylamine	0.400	0.419	-4.7	163	0.00
4 S	2-Fluorophenol	0.400	0.391	2.3	149	0.00
5 S	Phenol-d6	0.400	0.415	-3.7	160	0.00
6	bis(2-Chloroethyl)ether	0.400	0.510	-27.5#	212	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	148	0.00
8 S	Nitrobenzene-d5	0.400	0.423	-5.7	160	-0.01
9	Naphthalene	0.400	0.427	-6.7	159	0.00
10	Hexachlorobutadiene	0.400	0.385	3.8	134	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.436	-9.0	153	0.00
12	2-Methylnaphthalene	0.400	0.392	2.0	149	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	138	0.00
14 S	2,4,6-Tribromophenol	0.400	0.349	12.8	122	0.00
15 S	2-Fluorobiphenyl	0.400	0.467	-16.8	160	0.00
16	Acenaphthylene	0.400	0.470	-17.5	167	0.00
17	Acenaphthene	0.400	0.420	-5.0	148	0.00
18	Fluorene	0.400	0.421	-5.2	148	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	130	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.360	10.0	154	0.00
21	4-Bromophenyl-phenylether	0.400	0.402	-0.5	138	0.00
22	Hexachlorobenzene	0.400	0.434	-8.5	135	0.00
23	Atrazine	0.400	0.436	-9.0	150	0.00
24	Pentachlorophenol	0.400	0.347	13.3	127	0.00
25	Phenanthrene	0.400	0.428	-7.0	146	0.00
26	Anthracene	0.400	0.431	-7.7	149	0.00
27 SURR	Fluoranthene-d10	0.400	0.409	-2.2	128	0.00
28	Fluoranthene	0.400	0.392	2.0	134	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	129	0.00
30	Pyrene	0.400	0.439	-9.7	134	0.00
31 S	Terphenyl-d14	0.400	0.445	-11.2	137	0.00
32	Benzo(a)anthracene	0.400	0.430	-7.5	149	0.00
33	Chrysene	0.400	0.403	-0.8	129	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.429	-7.2	136	0.00
35 I	Perylene-d12	0.400	0.400	0.0	125	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.429	-7.2	141	0.00
37	Benzo(b)fluoranthene	0.400	0.390	2.5	132	0.00
38	Benzo(k)fluoranthene	0.400	0.430	-7.5	136	0.00
39 C	Benzo(a)pyrene	0.400	0.451	-12.7	147	0.00
40	Dibenzo(a,h)anthracene	0.400	0.404	-1.0	138	0.00
41	Benzo(g,h,i)perylene	0.400	0.412	-3.0	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN061925