

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062022\
 Data File : BN020276.D
 Acq On : 20 Jun 2022 13:40
 Operator : CG/JU
 Sample : N3294-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4P6

Quant Time: Jun 21 02:37:45 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN061822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 20 01:13:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.796	152	895	0.400	ng/ul	0.00
4) Naphthalene-d8	10.579	136	3580	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.419	164	3884	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.162	188	7699	0.400	ng/ul #	0.00
17) Chrysene-d12	21.355	240	10527	0.400	ng/ul #	0.00
23) Perylene-d12	23.667	264	12209	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.256	96	4192	4.102	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.174	152	2001	0.338	ng/ul	0.00
18) Fluoranthene-d10	19.192	212	8749	0.403	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.290	88	981	0.997	ng/ul#	28
5) Naphthalene	10.628	128	477	0.042	ng/ul#	13
7) 2-Methylnaphthalene	12.245	142	269	0.036	ng/ul	86
8) 1-Methylnaphthalene	12.465	142	294	0.040	ng/ul	91
15) Phenanthrene	17.200	178	684	0.025	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062022\
Data File : BN020276.D
Acq On : 20 Jun 2022 13:40
Operator : CG/JU
Sample : N3294-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4P6

Quant Time: Jun 21 02:37:45 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN061822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 20 01:13:51 2022
Response via : Initial Calibration

