


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062022\
 Data File : BN020270.D
 Acq On : 20 Jun 2022 09:52
 Operator : CG/JU
 Sample : DFTPP256
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP256

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN061822.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 20 01:13:51 2022



AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.8	35709	PASS
68	69	0.00	2	1.6	503	PASS
69	198	0.00	100	38.4	32005	PASS
70	69	0.00	2	0.5	153	PASS
127	198	10	80	49.7	41448	PASS
197	198	0.00	2	0.2	157	PASS
198	198	100	100	100.0	83397	PASS
199	198	5	9	6.8	5682	PASS
275	198	10	60	24.5	20429	PASS
365	198	1	100	3.0	2468	PASS
441	198	0.01	100	12.6	10549	PASS
442	198	50	100	70.7	58997	PASS
443	442	15	24	19.4	11471	PASS

m/z	Abundance
39.90	831.0
43.00	252.0
43.90	1184.0
48.90	381.0
72.90	296.0
84.00	302.0
85.80	160.0
193.10	201.0
207.00	778.0
208.00	216.0
281.00	179.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

m/z	Abundance
37.10	160.0
38.00	432.0
39.00	3153.0
40.00	1040.0
41.00	274.0
43.90	1030.0
49.00	573.0
50.00	10297.0
51.00	43024.0
52.00	1926.0
54.90	335.0
56.00	1308.0
56.90	2727.0
61.00	611.0
62.00	568.0
63.00	1617.0
64.10	294.0
65.10	744.0
68.10	707.0
69.00	37952.0
70.00	262.0
72.90	708.0
74.00	4458.0
75.00	5663.0
76.10	2224.0
77.00	42904.0
78.00	2990.0
78.90	3043.0
80.00	2374.0
81.00	2937.0
82.00	1011.0
83.00	631.0
83.80	283.0
85.00	432.0
85.90	1029.0
87.00	459.0
87.90	153.0
91.00	801.0
91.90	765.0
93.00	5756.0
94.00	365.0
95.00	151.0
96.00	299.0
97.20	223.0
97.90	3922.0
99.00	4216.0
99.90	450.0
101.00	1814.0
102.90	513.0
103.90	958.0
104.90	1143.0
106.00	533.0
107.00	13404.0
108.00	2038.0
108.90	396.0
110.00	25128.0
110.90	3561.0
112.00	417.0
112.90	240.0
116.00	807.0
117.00	10899.0
117.90	1057.0
121.90	931.0
122.90	1316.0
123.90	800.0
125.00	742.0
127.00	47184.0
128.00	3444.0
128.90	17496.0
130.00	1438.0
130.90	323.0
132.80	160.0
134.00	598.0
135.00	1290.0
136.00	632.0
137.00	1307.0
137.60	166.0
141.00	2131.0
141.90	851.0
142.80	550.0
144.00	164.0
146.00	398.0
147.00	1335.0
147.90	2849.0
149.00	821.0
149.80	207.0
150.80	301.0
151.40	277.0
152.00	213.0
152.90	665.0
153.90	601.0
154.90	1366.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

156.00	1931.0
156.90	477.0
157.80	347.0
158.90	224.0
159.90	715.0
160.90	1375.0
162.00	345.0
162.80	181.0
164.90	822.0
165.90	754.0
167.00	4611.0
167.90	4576.0
168.90	768.0
170.00	206.0
171.90	496.0
172.90	564.0
173.90	958.0
175.00	1841.0
176.10	556.0
176.90	824.0
178.00	339.0
178.90	2998.0
180.00	3035.0
181.00	1186.0
181.90	161.0
183.90	253.0
184.90	1620.0
186.00	12281.0
187.00	3519.0
188.00	600.0
188.90	534.0
190.90	466.0
191.90	903.0
192.90	1075.0
194.00	331.0
196.00	2418.0
197.00	470.0
198.00	87424.0
198.90	5873.0
199.90	514.0
200.60	195.0
201.40	536.0
202.90	650.0
204.00	3381.0
205.00	5047.0
206.00	22224.0
207.00	4222.0
207.90	697.0
208.90	417.0
209.90	242.0
210.90	1022.0
211.80	192.0
214.80	401.0
216.00	470.0
216.90	6215.0
217.90	764.0
221.00	3225.0
221.80	905.0
222.90	1412.0
224.00	12151.0
225.00	3251.0
226.20	468.0
226.90	5171.0
227.90	698.0
229.00	1118.0
230.80	399.0
233.80	421.0
234.90	374.0
235.90	181.0
237.00	791.0
238.00	173.0
239.80	222.0
240.90	256.0
242.00	599.0
243.00	734.0
244.00	8271.0
244.90	1068.0
245.90	1577.0
247.00	318.0
249.00	451.0
253.00	166.0
254.00	437.0
255.00	43824.0
256.00	6388.0
256.90	839.0
257.90	2388.0
258.90	325.0
265.00	1008.0
273.00	1534.0
274.00	3723.0
275.00	19200.0
276.00	2525.0
277.00	1788.0
277.90	437.0
281.00	291.0
282.90	213.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

284.90	257.0
292.90	406.0
293.90	246.0
295.90	5565.0
296.90	685.0
303.00	459.0
314.00	349.0
315.00	681.0
315.80	234.0
320.90	238.0
323.10	1523.0
324.00	338.0
326.90	326.0
333.90	973.0
334.90	255.0
340.90	238.0
345.90	380.0
352.00	606.0
353.00	496.0
353.90	544.0
354.90	194.0
365.00	2177.0
366.00	278.0
371.90	923.0
373.00	224.0
383.00	244.0
401.90	307.0
402.90	386.0
420.90	370.0
422.00	454.0
422.90	3017.0
423.90	1636.0
425.10	150.0
441.00	8320.0
442.00	45024.0
443.00	9895.0
444.00	868.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

m/z	Abundance
36.90	303.0
38.00	528.0
39.00	2968.0
39.90	988.0
41.00	328.0
43.10	305.0
44.00	1263.0
48.90	633.0
50.00	10325.0
51.00	42536.0
52.00	2321.0
55.00	345.0
55.90	1426.0
57.00	2770.0
61.00	533.0
62.00	612.0
63.00	1863.0
63.90	290.0
65.00	796.0
68.00	801.0
69.00	37328.0
70.00	196.0
72.90	698.0
74.00	3720.0
75.00	6604.0
76.00	2246.0
77.00	43560.0
78.00	3211.0
79.00	3198.0
79.90	2098.0
80.90	3230.0
81.90	724.0
83.00	760.0
83.90	307.0
84.90	621.0
86.00	1036.0
86.90	480.0
87.90	158.0
90.90	658.0
91.90	826.0
92.90	5199.0
93.80	352.0
95.90	287.0
97.90	4079.0
99.00	4002.0
100.00	382.0
101.00	1879.0
103.00	621.0
103.90	1145.0
104.90	1309.0
106.00	344.0
107.00	13012.0
108.00	2300.0
109.10	467.0
110.00	26488.0
111.00	4094.0
111.90	482.0
112.90	199.0
116.00	814.0
117.00	11541.0
117.80	1152.0
118.80	238.0
120.00	263.0
122.00	1063.0
123.00	1498.0
123.90	692.0
125.00	693.0
127.00	48368.0
128.00	3861.0
128.90	19440.0
130.00	1834.0
130.90	275.0
131.90	218.0
134.00	579.0
134.90	1656.0
135.90	545.0
137.00	1035.0
137.90	213.0
139.90	415.0
140.90	2371.0
142.00	639.0
143.00	780.0
143.90	203.0
145.80	504.0
146.90	1564.0
148.00	2770.0
148.90	645.0
150.00	202.0
151.00	287.0
151.80	163.0
152.90	823.0
154.00	637.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

155.00	1287.0
156.00	1940.0
157.10	406.0
157.90	532.0
158.80	289.0
160.00	753.0
160.90	988.0
162.00	482.0
164.80	1087.0
166.00	718.0
167.00	4965.0
168.00	4426.0
169.00	719.0
171.90	397.0
172.90	586.0
174.00	946.0
174.90	1933.0
176.10	535.0
176.90	851.0
177.80	349.0
178.90	3471.0
180.00	3228.0
181.00	1185.0
181.90	268.0
183.90	211.0
185.00	1864.0
186.00	13878.0
187.00	3989.0
188.00	521.0
188.90	929.0
190.10	210.0
190.90	476.0
191.90	949.0
192.90	1409.0
194.10	277.0
195.90	3023.0
197.90	98248.0
198.90	6750.0
200.00	614.0
201.30	361.0
202.90	584.0
204.00	3984.0
205.00	5674.0
206.00	26016.0
207.00	4451.0
208.00	995.0
209.00	506.0
210.00	446.0
211.00	1026.0
215.00	270.0
215.90	514.0
216.90	6589.0
217.90	735.0
221.00	3985.0
221.80	868.0
223.00	1556.0
224.00	12704.0
225.00	3542.0
225.90	404.0
227.00	5438.0
227.90	684.0
228.90	968.0
230.90	558.0
233.90	358.0
234.90	357.0
235.80	265.0
237.00	754.0
239.00	241.0
241.00	291.0
242.00	547.0
242.90	668.0
244.00	10536.0
245.00	1343.0
246.00	1857.0
246.90	536.0
248.90	467.0
252.90	204.0
254.10	412.0
255.00	55344.0
256.00	7888.0
256.90	928.0
258.00	3064.0
258.90	646.0
265.00	1419.0
266.00	200.0
271.90	225.0
273.00	1704.0
274.00	4498.0
275.00	25640.0
276.00	3468.0
277.00	1914.0
277.90	524.0
280.80	192.0
283.00	246.0
283.90	236.0
284.90	291.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

292.80	576.0
293.90	196.0
294.90	206.0
296.00	7287.0
297.00	834.0
303.00	677.0
304.00	208.0
314.00	270.0
314.90	573.0
315.90	502.0
320.90	193.0
323.00	2234.0
324.10	247.0
326.90	510.0
327.70	195.0
332.00	196.0
333.00	252.0
334.00	1273.0
334.80	408.0
341.00	281.0
345.90	438.0
352.00	810.0
353.00	585.0
353.90	821.0
354.90	179.0
364.90	2882.0
365.90	455.0
371.00	244.0
372.00	1184.0
373.10	274.0
383.00	276.0
401.90	283.0
402.90	636.0
403.80	238.0
420.90	482.0
422.00	559.0
423.00	4778.0
424.00	1139.0
425.10	223.0
441.00	13617.0
442.00	68648.0
443.00	14151.0
444.00	1460.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

m/z	Abundance
37.90	322.0
39.00	1788.0
39.90	1079.0
41.00	248.0
43.00	163.0
43.90	946.0
49.00	563.0
50.00	6003.0
51.00	21568.0
52.00	1146.0
55.00	282.0
55.90	805.0
56.90	1616.0
61.00	281.0
62.00	322.0
63.00	881.0
63.80	179.0
65.00	412.0
68.90	20736.0
72.90	644.0
73.90	2051.0
75.00	3572.0
76.00	1305.0
77.00	24480.0
78.00	1984.0
79.00	1738.0
80.00	1073.0
81.00	2003.0
82.00	589.0
83.00	437.0
83.90	222.0
85.10	388.0
85.90	602.0
86.80	230.0
90.90	410.0
91.90	539.0
92.90	2932.0
93.80	292.0
96.00	297.0
98.00	2661.0
99.00	2092.0
99.80	268.0
100.90	1146.0
103.00	451.0
104.00	523.0
105.00	523.0
106.10	215.0
107.00	8549.0
108.00	1495.0
108.80	221.0
110.00	14592.0
110.90	2435.0
111.90	304.0
115.90	498.0
117.00	6740.0
117.90	816.0
121.80	571.0
123.00	757.0
123.80	234.0
124.90	397.0
127.00	28792.0
128.00	2015.0
129.00	11235.0
129.90	869.0
131.00	247.0
133.90	322.0
134.90	1246.0
136.00	330.0
136.90	664.0
139.90	166.0
140.90	1488.0
142.00	484.0
142.80	304.0
145.80	223.0
147.00	883.0
148.00	1661.0
148.90	447.0
150.00	163.0
151.10	203.0
152.90	643.0
154.00	348.0
155.00	905.0
156.00	1127.0
156.90	193.0
157.90	352.0
159.00	244.0
160.00	440.0
161.00	725.0
162.00	225.0
164.90	522.0
165.80	442.0
167.00	2769.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

168.00	2666.0
168.90	280.0
171.00	175.0
171.90	317.0
173.00	431.0
174.00	736.0
174.90	1299.0
175.90	252.0
177.00	676.0
177.90	285.0
178.90	2053.0
180.00	2090.0
181.00	629.0
182.00	163.0
184.00	264.0
185.00	1346.0
185.90	8074.0
187.00	2426.0
188.10	384.0
188.90	661.0
190.90	275.0
192.00	771.0
193.00	901.0
195.90	1719.0
197.90	64520.0
198.90	4424.0
199.80	244.0
200.20	179.0
201.30	313.0
201.70	306.0
203.00	390.0
203.90	2308.0
204.90	3899.0
206.00	16059.0
207.00	3165.0
208.00	709.0
208.90	306.0
209.80	316.0
210.90	607.0
211.80	182.0
214.90	168.0
215.90	331.0
216.90	4480.0
218.00	513.0
220.90	2210.0
221.80	626.0
222.90	947.0
224.00	8723.0
225.00	2283.0
226.00	202.0
227.00	3636.0
227.80	522.0
228.90	753.0
230.00	250.0
230.80	296.0
234.00	202.0
234.90	284.0
235.90	194.0
236.90	500.0
241.00	175.0
241.90	653.0
243.00	488.0
244.00	7166.0
245.00	903.0
246.00	1217.0
246.90	347.0
255.00	36024.0
256.00	5433.0
256.90	340.0
257.90	2233.0
259.00	338.0
264.90	979.0
265.90	159.0
273.00	1308.0
273.90	2899.0
275.00	16448.0
276.00	2265.0
277.00	1597.0
277.90	388.0
281.00	206.0
283.00	270.0
285.00	265.0
292.90	506.0
296.00	4773.0
296.90	661.0
303.00	576.0
303.90	182.0
313.80	286.0
315.00	412.0
316.00	369.0
322.90	1566.0
324.00	280.0
326.90	280.0
327.90	173.0
333.00	174.0
333.90	1091.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256

334.90	249.0
341.00	242.0
345.90	336.0
352.00	443.0
352.90	364.0
354.00	644.0
365.00	2346.0
365.80	208.0
371.90	860.0
372.90	302.0
382.90	283.0
389.90	154.0
402.00	435.0
402.90	576.0
403.80	183.0
421.00	433.0
421.90	459.0
422.90	3548.0
424.00	1239.0
425.10	164.0
441.00	9709.0
442.00	63320.0
442.90	10367.0
443.90	1139.0

Instrument :
BNA_N
ClientSampleId :
DFTPP256