

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062125\
 Data File : BN037362.D
 Acq On : 20 Jun 2025 22:52
 Operator : RC/JU
 Sample : Q2377-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PW-B6-L66-061925

Quant Time: Jun 20 23:51:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 20 23:41:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.568	152	1844	0.400	ng	0.00
7) Naphthalene-d8	10.351	136	4234	0.400	ng	# 0.01
13) Acenaphthene-d10	14.213	164	2932	0.400	ng	0.00
19) Phenanthrene-d10	16.971	188	6082	0.400	ng	# 0.00
29) Chrysene-d12	21.162	240	5238	0.400	ng	0.00
35) Perylene-d12	23.348	264	5349	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.170	112	526	0.143	ng	0.00
5) Phenol-d6	6.759	99	262	0.069	ng	0.00
8) Nitrobenzene-d5	8.717	82	1049	0.307	ng	0.00
11) 2-Methylnaphthalene-d10	11.945	152	2285	0.333	ng	0.00
14) 2,4,6-Tribromophenol	15.718	330	628	0.365	ng	0.00
15) 2-Fluorobiphenyl	12.838	172	4543	0.353	ng	0.00
27) Fluoranthene-d10	19.003	212	6541	0.375	ng	0.00
31) Terphenyl-d14	19.612	244	5026	0.421	ng	0.00
Target Compounds						Qvalue
23) Atrazine	16.450	200	219	0.063	ng	# 91
34) Bis(2-ethylhexyl)phtha...	21.090	149	1173	0.166	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062125\
Data File : BN037362.D
Acq On : 20 Jun 2025 22:52
Operator : RC/JU
Sample : Q2377-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PW-B6-L66-061925

Quant Time: Jun 20 23:51:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 20 23:41:54 2025
Response via : Initial Calibration

