

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
 Data File : BN011015.D
 Acq On : 22 Jun 2020 18:39
 Operator : CG/JU
 Sample : L3033-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 KY032MW006-200616

Manual Integrations
 APPROVED

Sohil
 6/23/2020 2:12:22 PM

Quant Time: Jun 22 19:16:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 22 12:02:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.53	152	4523	0.40	ng	0.00
7) Naphthalene-d8	10.25	136	33140m	0.40	ng	-0.03
13) Acenaphthene-d10	14.16	164	9935	0.40	ng	0.01
19) Phenanthrene-d10	16.90	188	12782	0.40	ng	-0.01
27) Chrysene-d12	21.11	240	10719	0.40	ng	-0.01
34) Perylene-d12	23.26	264	13945	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.20	112	2087	0.24	ng	0.02
5) Phenol-d6	6.74	99	4873	0.47	ng	0.02
8) Nitrobenzene-d5	8.67	82	4088m	0.17	ng	0.00
11) 2-Methylnaphthalene-d10	11.88	152	7983m	0.16	ng	0.00
14) 2,4,6-Tribromophenol	15.66	330	1412	0.40	ng	0.00
15) 2-Fluorobiphenyl	12.78	172	11000	0.28	ng	0.00
25) Fluoranthene-d10	18.94	212	13106	0.37	ng	0.00
29) Terphenyl-d14	19.56	244	13811	0.55	ng	0.00

Target Compounds						Qvalue
2) 1,4-Dioxane	3.22	88	10161224	2293.573	ng	98
6) bis(2-Chloroethyl)ether	6.99	93	2946940	293.855	ng	97
9) Naphthalene	10.34	128	1279373	15.362	ng	97
12) 2-Methylnaphthalene	11.95	142	118529	2.127	ng	# 96
17) Acenaphthene	14.20	154	1653	0.060	ng	# 65
18) Fluorene	15.21	166	2384	0.067	ng	# 46
23) Phenanthrene	16.94	178	1774	0.051	ng	# 60
24) Anthracene	17.04	178	1324	0.045	ng	# 21
26) Fluoranthene	18.97	202	1050	0.026	ng	# 92
32) Bis(2-ethylhexyl)phthalate	21.06	149	6673	0.635	ng	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
Data File : BN011015.D
Acq On : 22 Jun 2020 18:39
Operator : CG/JU
Sample : L3033-06
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
KY032MW006-200616

Manual Integrations
APPROVED

Sohil
6/23/2020 2:12:22 PM

Quant Time: Jun 22 19:16:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 22 12:02:11 2020
Response via : Initial Calibration

