

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
 Data File : BN011002.D
 Acq On : 22 Jun 2020 10:53
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 22 12:03:54 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 22 12:02:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.392	0.470	-19.9	121	0.00
3	n-Nitrosodimethylamine	0.228	0.248	-8.8	131	0.00
4 S	2-Fluorophenol	0.772	0.886	-14.8	119	0.00
5 S	Phenol-d6	0.910	0.981	-7.8	114	0.00
6	bis(2-Chloroethyl)ether	0.887	0.971	-9.5	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.296	0.307	-3.7	101	0.00
9	Naphthalene	1.005	1.051	-4.6	101	0.00
10	Hexachlorobutadiene	0.248	0.276	-11.3	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.618	0.3	96	0.00
12	2-Methylnaphthalene	0.673	0.667	0.9	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.143	0.142	0.7	100	0.00
15 S	2-Fluorobiphenyl	1.578	1.655	-4.9	96	0.00
16	Acenaphthylene	1.639	1.703	-3.9	100	0.00
17	Acenaphthene	1.117	1.146	-2.6	96	0.00
18	Fluorene	1.441	1.433	0.6	94	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.240	0.268	-11.7	104	0.00
21	Hexachlorobenzene	0.251	0.290	-15.5	105	0.00
22	Pentachlorophenol	0.084	0.077	8.3	90	0.00
23	Phenanthrene	1.093	1.098	-0.5	93	0.00
24	Anthracene	0.912	0.903	1.0	95	0.00
25 SURR	Fluoranthene-d10	1.104	1.121	-1.5	96	0.00
26	Fluoranthene	1.240	1.239	0.1	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
28	Pvrene	1.323	1.296	2.0	96	0.00
29 S	Terphenyl-d14	0.936	0.931	0.5	96	0.00
30	Benzo(a)anthracene	1.176	1.186	-0.9	103	0.00
31	Chrysene	1.370	1.389	-1.4	95	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.448	-14.3	119	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.684	-10.0	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.360	1.324	2.6	106	0.00
36	Benzo(k)fluoranthene	1.449	1.390	4.1	102	0.00
37 C	Benzo(a)pyrene	1.167	1.177	-0.9	113	0.00
38	Dibenzo(a,h)anthracene	1.330	1.301	2.2	108	0.00
39	Benzo(a,h,i)perylene	1.391	1.385	0.4	105	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
Data File : BN011002.D
Acq On : 22 Jun 2020 10:53
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 22 12:03:54 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 22 12:02:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			