

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
 Data File : BN011018.D
 Acq On : 22 Jun 2020 20:27
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 23 06:27:04 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 22 12:02:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	137	0.00
2	1,4-Dioxane	0.392	0.483	-23.2	162#	0.00
3	n-Nitrosodimethylamine	0.228	0.275	-20.6	188#	-0.03
4 S	2-Fluorophenol	0.772	1.048	-35.8#	182#	0.00
5 S	Phenol-d6	0.910	1.274	-40.0#	192#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.220	-37.5#	189#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	154#	0.00
8 S	Nitrobenzene-d5	0.296	0.323	-9.1	161#	-0.01
9	Naphthalene	1.005	1.034	-2.9	151#	0.00
10	Hexachlorobutadiene	0.248	0.244	1.6	140	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.658	-6.1	155#	0.00
12	2-Methylnaphthalene	0.673	0.714	-6.1	159#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	163#	0.00
14 S	2,4,6-Tribromophenol	0.143	0.154	-7.7	181#	-0.01
15 S	2-Fluorobiphenyl	1.578	1.570	0.5	153#	-0.01
16	Acenaphthylene	1.639	1.731	-5.6	171#	0.00
17	Acenaphthene	1.117	1.152	-3.1	162#	0.00
18	Fluorene	1.441	1.432	0.6	158#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	157#	-0.01
20	4-Bromophenyl-phenylether	0.240	0.238	0.8	152#	-0.01
21	Hexachlorobenzene	0.251	0.250	0.4	149	0.00
22	Pentachlorophenol	0.084	0.087	-3.6	168#	0.00
23	Phenanthrene	1.093	1.103	-0.9	155#	0.00
24	Anthracene	0.912	1.021	-12.0	177#	0.00
25 SURR	Fluoranthene-d10	1.104	1.161	-5.2	163#	0.00
26	Fluoranthene	1.240	1.274	-2.7	161#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	161#	-0.01
28	Pvrene	1.323	1.417	-7.1	166#	0.00
29 S	Terphenyl-d14	0.936	1.005	-7.4	165#	0.00
30	Benzo(a)anthracene	1.176	1.321	-12.3	182#	0.00
31	Chrysene	1.370	1.352	1.3	147	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.740	-88.8#	310#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.615	-5.5	162#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	178#	0.00
35	Benzo(b)fluoranthene	1.360	1.296	4.7	164#	-0.01
36	Benzo(k)fluoranthene	1.449	1.290	11.0	151#	0.00
37 C	Benzo(a)pyrene	1.167	1.188	-1.8	181#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.249	6.1	165#	-0.01
39	Benzo(a,h,i)perylene	1.391	1.264	9.1	153#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
Data File : BN011018.D
Acq On : 22 Jun 2020 20:27
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 23 06:27:04 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 22 12:02:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			