

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
 Data File : BN011020.D
 Acq On : 23 Jun 2020 03:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 23 05:34:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 22 12:02:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	145	0.00
2	1,4-Dioxane	0.392	0.450	-14.8	160#	0.00
3	n-Nitrosodimethylamine	0.228	0.268	-17.5	194#	-0.02
4 S	2-Fluorophenol	0.772	0.978	-26.7#	180#	0.00
5 S	Phenol-d6	0.910	1.030	-13.2	164#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.194	-34.6#	196#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	142	0.00
8 S	Nitrobenzene-d5	0.296	0.313	-5.7	144	-0.01
9	Naphthalene	1.005	1.047	-4.2	141	-0.01
10	Hexachlorobutadiene	0.248	0.244	1.6	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.646	-4.2	141	0.00
12	2-Methylnaphthalene	0.673	0.703	-4.5	144	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	144	0.00
14 S	2,4,6-Tribromophenol	0.143	0.143	0.0	149	-0.01
15 S	2-Fluorobiphenyl	1.578	1.648	-4.4	141	-0.01
16	Acenaphthylene	1.639	1.736	-5.9	151#	0.00
17	Acenaphthene	1.117	1.168	-4.6	145	0.00
18	Fluorene	1.441	1.500	-4.1	146	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	146	-0.01
20	4-Bromophenyl-phenylether	0.240	0.268	-11.7	160#	-0.01
21	Hexachlorobenzene	0.251	0.276	-10.0	154#	0.00
22	Pentachlorophenol	0.084	0.073	13.1	131	0.00
23	Phenanthrene	1.093	1.106	-1.2	145	0.00
24	Anthracene	0.912	0.944	-3.5	153#	0.00
25 SURR	Fluoranthene-d10	1.104	1.132	-2.5	149	0.00
26	Fluoranthene	1.240	1.269	-2.3	150	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	158#	-0.01
28	Pvrene	1.323	1.320	0.2	152#	0.00
29 S	Terphenyl-d14	0.936	0.933	0.3	150#	0.00
30	Benzo(a)anthracene	1.176	1.215	-3.3	165#	-0.01
31	Chrysene	1.370	1.363	0.5	145	-0.01
32	Bis(2-ethylhexyl)phthalate	0.392	0.487	-24.2	201#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.531	1.706	-11.4	168#	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	176#	-0.01
35	Benzo(b)fluoranthene	1.360	1.295	4.8	162#	-0.01
36	Benzo(k)fluoranthene	1.449	1.383	4.6	160#	-0.01
37 C	Benzo(a)pyrene	1.167	1.142	2.1	172#	-0.01
38	Dibenzo(a,h)anthracene	1.330	1.291	2.9	168#	-0.02
39	Benzo(a,h,i)perylene	1.391	1.478	-6.3	176#	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
Data File : BN011020.D
Acq On : 23 Jun 2020 03:28
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 23 05:34:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 22 12:02:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			