

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
 Data File : BN011020.D
 Acq On : 23 Jun 2020 03:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 23 05:34:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 22 12:02:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	145	0.00
2	1,4-Dioxane	0.400	0.460	-15.0	160	0.00
3	n-Nitrosodimethylamine	0.400	0.471	-17.7	194	-0.02
4 S	2-Fluorophenol	0.400	0.507	-26.7#	180	0.00
5 S	Phenol-d6	0.400	0.453	-13.2	164	0.00
6	bis(2-Chloroethyl)ether	0.400	0.539	-34.8#	196	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	142	0.00
8 S	Nitrobenzene-d5	0.400	0.423	-5.7	144	-0.01
9	Naphthalene	0.400	0.417	-4.2	141	-0.01
10	Hexachlorobutadiene	0.400	0.393	1.8	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.417	-4.2	141	0.00
12	2-Methylnaphthalene	0.400	0.418	-4.5	144	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	144	0.00
14 S	2,4,6-Tribromophenol	0.400	0.400	0.0	149	-0.01
15 S	2-Fluorobiphenyl	0.400	0.418	-4.5	141	-0.01
16	Acenaphthylene	0.400	0.424	-6.0	151	0.00
17	Acenaphthene	0.400	0.418	-4.5	145	0.00
18	Fluorene	0.400	0.416	-4.0	146	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	146	-0.01
20	4-Bromophenyl-phenylether	0.400	0.447	-11.7	160	-0.01
21	Hexachlorobenzene	0.400	0.441	-10.2	154	0.00
22	Pentachlorophenol	0.400	0.345	13.8	131	0.00
23	Phenanthrene	0.400	0.405	-1.3	145	0.00
24	Anthracene	0.400	0.414	-3.5	153	0.00
25 SURR	Fluoranthene-d10	0.400	0.410	-2.5	149	0.00
26	Fluoranthene	0.400	0.409	-2.2	150	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	158	-0.01
28	Pvrene	0.400	0.399	0.3	152	0.00
29 S	Terphenyl-d14	0.400	0.398	0.5	150	0.00
30	Benzo(a)anthracene	0.400	0.414	-3.5	165	-0.01
31	Chrysene	0.400	0.398	0.5	145	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.496	-24.0	201	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.446	-11.5	168	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	176	-0.01
35	Benzo(b)fluoranthene	0.400	0.381	4.8	162	-0.01
36	Benzo(k)fluoranthene	0.400	0.382	4.5	160	-0.01
37 C	Benzo(a)pyrene	0.400	0.392	2.0	172	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.388	3.0	168	-0.02
39	Benzo(a,h,i)perylene	0.400	0.425	-6.2	176	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
Data File : BN011020.D
Acq On : 23 Jun 2020 03:28
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 23 05:34:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 22 12:02:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				