

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
 Data File : BN011028.D
 Acq On : 23 Jun 2020 08:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 23 08:46:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 22 12:02:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	140	0.00
2	1,4-Dioxane	0.392	0.501	-27.8#	171#	0.00
3	n-Nitrosodimethylamine	0.228	0.292	-28.1#	204#	-0.02
4 S	2-Fluorophenol	0.772	1.059	-37.2#	188#	0.00
5 S	Phenol-d6	0.910	1.172	-28.8#	180#	0.00
6	bis(2-Chloroethyl)ether	0.887	1.100	-24.0	174#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	150#	0.00
8 S	Nitrobenzene-d5	0.296	0.343	-15.9	167#	-0.01
9	Naphthalene	1.005	1.062	-5.7	151#	-0.01
10	Hexachlorobutadiene	0.248	0.242	2.4	136	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.661	-6.6	153#	0.00
12	2-Methylnaphthalene	0.673	0.721	-7.1	157#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	157#	0.00
14 S	2,4,6-Tribromophenol	0.143	0.142	0.7	161#	-0.01
15 S	2-Fluorobiphenyl	1.578	1.613	-2.2	150#	-0.01
16	Acenaphthylene	1.639	1.766	-7.7	167#	-0.01
17	Acenaphthene	1.117	1.161	-3.9	156#	0.00
18	Fluorene	1.441	1.472	-2.2	156#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	178#	-0.01
20	4-Bromophenyl-phenylether	0.240	0.202	15.8	147	-0.01
21	Hexachlorobenzene	0.251	0.219	12.7	148	0.00
22	Pentachlorophenol	0.084	0.079	6.0	172#	0.00
23	Phenanthrene	1.093	1.095	-0.2	175#	0.00
24	Anthracene	0.912	0.941	-3.2	186#	0.00
25 SURR	Fluoranthene-d10	1.104	1.035	6.3	166#	0.00
26	Fluoranthene	1.240	1.152	7.1	166#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	172#	-0.01
28	Pvrene	1.323	1.348	-1.9	169#	0.00
29 S	Terphenyl-d14	0.936	0.938	-0.2	165#	0.00
30	Benzo(a)anthracene	1.176	1.237	-5.2	183#	0.00
31	Chrysene	1.370	1.411	-3.0	164#	0.00
32	Bis(2-ethylhexyl)phthalate	0.392	0.539	-37.5#	242#	0.00
33	Indeno(1,2,3-cd)pyrene	1.531	1.715	-12.0	185#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	219#	0.00
35	Benzo(b)fluoranthene	1.360	1.141	16.1	178#	0.00
36	Benzo(k)fluoranthene	1.449	1.242	14.3	179#	0.00
37 C	Benzo(a)pyrene	1.167	1.237	-6.0	231#	0.00
38	Dibenzo(a,h)anthracene	1.330	1.143	14.1	185#	-0.01
39	Benzo(a,h,i)perylene	1.391	1.142	17.9	169#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
Data File : BN011028.D
Acq On : 23 Jun 2020 08:15
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 23 08:46:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 22 12:02:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			