

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
 Data File : BN011028.D
 Acq On : 23 Jun 2020 08:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 23 08:46:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 22 12:02:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	140	0.00
2	1,4-Dioxane	0.400	0.512	-28.0#	171	0.00
3	n-Nitrosodimethylamine	0.400	0.513	-28.2#	204	-0.02
4 S	2-Fluorophenol	0.400	0.549	-37.3#	188	0.00
5 S	Phenol-d6	0.400	0.516	-29.0#	180	0.00
6	bis(2-Chloroethyl)ether	0.400	0.496	-24.0	174	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	150	0.00
8 S	Nitrobenzene-d5	0.400	0.463	-15.8	167	-0.01
9	Naphthalene	0.400	0.423	-5.7	151	-0.01
10	Hexachlorobutadiene	0.400	0.390	2.5	136	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.426	-6.5	153	0.00
12	2-Methylnaphthalene	0.400	0.429	-7.2	157	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	157	0.00
14 S	2,4,6-Tribromophenol	0.400	0.398	0.5	161	-0.01
15 S	2-Fluorobiphenyl	0.400	0.409	-2.2	150	-0.01
16	Acenaphthylene	0.400	0.431	-7.7	167	-0.01
17	Acenaphthene	0.400	0.416	-4.0	156	0.00
18	Fluorene	0.400	0.409	-2.2	156	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	178	-0.01
20	4-Bromophenyl-phenylether	0.400	0.337	15.8	147	-0.01
21	Hexachlorobenzene	0.400	0.349	12.8	148	0.00
22	Pentachlorophenol	0.400	0.373	6.8	172	0.00
23	Phenanthrene	0.400	0.401	-0.3	175	0.00
24	Anthracene	0.400	0.412	-3.0	186	0.00
25 SURR	Fluoranthene-d10	0.400	0.375	6.3	166	0.00
26	Fluoranthene	0.400	0.371	7.3	166	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	172	-0.01
28	Pvrene	0.400	0.407	-1.7	169	0.00
29 S	Terphenyl-d14	0.400	0.401	-0.3	165	0.00
30	Benzo(a)anthracene	0.400	0.421	-5.2	183	0.00
31	Chrysene	0.400	0.412	-3.0	164	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.550	-37.5#	242	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.448	-12.0	185	0.00
34 I	Perylene-d12	0.400	0.400	0.0	219	0.00
35	Benzo(b)fluoranthene	0.400	0.335	16.3	178	0.00
36	Benzo(k)fluoranthene	0.400	0.343	14.2	179	0.00
37 C	Benzo(a)pyrene	0.400	0.424	-6.0	231	0.00
38	Dibenzo(a,h)anthracene	0.400	0.344	14.0	185	-0.01
39	Benzo(a,h,i)perylene	0.400	0.328	18.0	169	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062220\
Data File : BN011028.D
Acq On : 23 Jun 2020 08:15
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 23 08:46:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN061020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 22 12:02:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				