

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062224\
 Data File : BN032182.D
 Acq On : 23 Jun 2024 01:08
 Operator : MA/JU
 Sample : P2899-08
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AB0

Quant Time: Jun 23 01:46:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 20 10:49:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	344260	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	1386587	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	776858	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.033	188	1362908	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.263	240	1123024	20.000	ng/u1	0.00
88) Perylene-d12	24.180	264	1374730	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.164	96	21289	2.558	ng/uL	0.00
4) Pyridine-d5	3.569	84	283599	12.107	ng/u1	0.00
7) Phenol-d5	6.816	99	456925	15.482	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.975	67	247983	14.365	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	401008	17.808	ng/u1	0.00
15) 4-Methylphenol-d8	8.346	113	363353	15.104	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	191807	18.092	ng/u1	0.00
24) 2-Nitrophenol-d4	9.504	143	212371	19.392	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.046	165	349180	17.079	ng/u1	0.00
31) 4-Chloroaniline-d4	10.563	131	460547	15.171	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	975372	18.022	ng/u1	0.00
49) Acenaphthylene-d8	13.969	160	1127115	18.112	ng/u1	0.00
54) 4-Nitrophenol-d4	14.522	143	135364	12.492	ng/u1	0.02
60) Fluorene-d10	15.275	176	791630	17.287	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.416	200	74016	10.804	ng/u1	0.00
73) Anthracene-d10	17.128	188	1096215	18.787	ng/u1	0.00
81) Pyrene-d10	19.433	212	1149040	19.120	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.980	264	1178130	17.947	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062224\
Data File : BN032182.D
Acq On : 23 Jun 2024 01:08
Operator : MA/JU
Sample : P2899-08
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AB0

Quant Time: Jun 23 01:46:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 20 10:49:40 2024
Response via : Initial Calibration

