

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN062320\
 Data File : BN011037.D
 Acq On : 23 Jun 2020 19:34
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062320

Quant Time: Jun 23 20:07:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 23 18:53:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.538	0.580	-7.8	100	0.00
3	n-Nitrosodimethylamine	0.300	0.295	1.7	92	0.00
4 S	2-Fluorophenol	1.127	1.145	-1.6	95	0.00
5 S	Phenol-d6	1.335	1.327	0.6	92	0.00
6	bis(2-Chloroethyl)ether	1.153	1.191	-3.3	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.364	0.359	1.4	81	0.01
9	Naphthalene	1.215	1.241	-2.1	98	0.00
10	Hexachlorobutadiene	0.293	0.316	-7.8	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.706	0.736	-4.2	92	0.00
12	2-Methylnaphthalene	0.814	0.853	-4.8	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.176	0.156	11.4	93	0.01
15 S	2-Fluorobiphenyl	1.866	2.004	-7.4	109	0.00
16	Acenaphthylene	2.016	1.993	1.1	105	0.00
17	Acenaphthene	1.354	1.369	-1.1	105	0.00
18	Fluorene	1.727	1.593	7.8	96	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4,6-Dinitro-2-methylphenol	0.067	0.060	10.4	91	0.01
21	4-Bromophenyl-phenylether	0.277	0.280	-1.1	94	0.01
22	Hexachlorobenzene	0.306	0.323	-5.6	96	0.01
23	Atrazine	0.204	0.201	1.5	96	0.00
24	Pentachlorophenol	0.097	0.090	7.2	88	0.00
25	Phenanthrene	1.348	1.397	-3.6	96	0.00
26	Anthracene	1.119	1.105	1.3	94	0.00
27 SURR	Fluoranthene-d10	1.265	1.252	1.0	94	0.00
28	Fluoranthene	1.541	1.525	1.0	91	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	105	-0.01
30	Pyrene	1.539	1.421	7.7	96	0.00
31 S	Terphenyl-d14	1.074	1.005	6.4	96	0.00
32	Benzo(a)anthracene	1.386	1.363	1.7	101	0.00
33	Chrysene	1.511	1.558	-3.1	108	-0.01
34	Bis(2-ethylhexyl)phthalate	0.557	0.506	9.2	94	0.00
35	Indeno(1,2,3-cd)pyrene	1.680	1.908	-13.6	113	-0.01
36 I	Perylene-d12	1.000	1.000	0.0	88	-0.01
37	Benzo(b)fluoranthene	1.582	1.800	-13.8	113	-0.01
38	Benzo(k)fluoranthene	1.645	2.008	-22.1	125	-0.01
39 C	Benzo(a)pyrene	1.405	1.422	-1.2	88	-0.01
40	Dibenzo(a,h)anthracene	1.515	1.692	-11.7	111	-0.01
41	Benzo(g,h,i)perylene	1.569	1.579	-0.6	103	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN062320\
Data File : BN011037.D
Acq On : 23 Jun 2020 19:34
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN062320

Quant Time: Jun 23 20:07:50 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 23 18:53:50 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0