

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
 Data File : BN014977.D
 Acq On : 22 Jun 2021 17:17
 Operator : CG/JU
 Sample : M2750-06
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D2-20210615

Manual Integrations
 APPROVED

mohammad
 6/23/2021 4:50:16 PM

Quant Time: Jun 22 17:59:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 01:26:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	19898	0.400	ng	#-0.01
7) Naphthalene-d8	10.518	136	88991	0.400	ng	0.00
13) Acenaphthene-d10	14.359	164	48130	0.400	ng	-0.01
19) Phenanthrene-d10	17.104	188	93256	0.400	ng	-0.01
29) Chrysene-d12	21.313	240	82730	0.400	ng	-0.01
35) Perylene-d12	23.564	264	72038	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.390	112	9075	0.147	ng	0.02
5) Phenol-d6	6.930	99	6310	0.081	ng	0.00
8) Nitrobenzene-d5	8.883	82	22766	0.339	ng	-0.02
11) 2-Methylnaphthalene-d10	12.106	152	50795	0.333	ng	-0.01
14) 2,4,6-Tribromophenol	15.856	330	8588	0.380	ng	-0.01
15) 2-Fluorobiphenyl	12.988	172	77542	0.424	ng	-0.01
27) Fluoranthene-d10	19.148	212	106358	0.362	ng	0.00
31) Terphenyl-d14	19.754	244	108927	0.531	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.361	88	104511m	12.523	ng	
34) Bis(2-ethylhexyl)phtha...	21.239	149	41974	0.193	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
Data File : BN014977.D
Acq On : 22 Jun 2021 17:17
Operator : CG/JU
Sample : M2750-06
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D2-20210615

Manual Integrations
APPROVED

mohammad
6/23/2021 4:50:16 PM

Quant Time: Jun 22 17:59:43 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 17 01:26:44 2021
Response via : Initial Calibration

