

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
 Data File : BN014979.D
 Acq On : 22 Jun 2021 18:29
 Operator : CG/JU
 Sample : M2750-08
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D1-20210615

Manual Integrations
 APPROVED

mohammad
 6/23/2021 4:50:24 PM

Quant Time: Jun 23 03:38:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 01:26:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	19924	0.400	ng	#-0.01
7) Naphthalene-d8	10.518	136	90178	0.400	ng	0.00
13) Acenaphthene-d10	14.359	164	48309	0.400	ng	-0.01
19) Phenanthrene-d10	17.104	188	93879	0.400	ng	-0.01
29) Chrysene-d12	21.303	240	81760	0.400	ng	#-0.02
35) Perylene-d12	23.561	264	69987	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.381	112	8132	0.132	ng	0.00
5) Phenol-d6	6.920	99	5765	0.074	ng	0.00
8) Nitrobenzene-d5	8.883	82	20278	0.298	ng	-0.02
11) 2-Methylnaphthalene-d10	12.106	152	44740	0.289	ng	-0.01
14) 2,4,6-Tribromophenol	15.856	330	5964	0.263	ng	-0.01
15) 2-Fluorobiphenyl	12.988	172	66576	0.363	ng	-0.01
27) Fluoranthene-d10	19.143	212	92560	0.313	ng	-0.01
31) Terphenyl-d14	19.754	244	99877	0.493	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.361	88	76560m	9.162	ng	Qvalue
6) bis(2-Chloroethyl)ether	7.188	93	1383	0.023	ng	88
34) Bis(2-ethylhexyl)phtha...	21.239	149	22610	0.105	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
Data File : BN014979.D
Acq On : 22 Jun 2021 18:29
Operator : CG/JU
Sample : M2750-08
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D1-20210615

Manual Integrations
APPROVED

mohammad
6/23/2021 4:50:24 PM

Quant Time: Jun 23 03:38:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 17 01:26:44 2021
Response via : Initial Calibration

