

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
 Data File : BN014982.D
 Acq On : 22 Jun 2021 20:16
 Operator : CG/JU
 Sample : M2750-11
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP01-20210615

Manual Integrations
 APPROVED

mohammad
 6/23/2021 4:50:31 PM

Quant Time: Jun 23 03:39:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 01:26:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	20104	0.400	ng	#-0.01
7) Naphthalene-d8	10.517	136	90666	0.400	ng	0.00
13) Acenaphthene-d10	14.358	164	47872	0.400	ng	-0.01
19) Phenanthrene-d10	17.103	188	95071	0.400	ng	-0.01
29) Chrysene-d12	21.302	240	82907	0.400	ng	#-0.02
35) Perylene-d12	23.560	264	71467	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.380	112	9147	0.147	ng	0.00
5) Phenol-d6	6.920	99	6279	0.080	ng	0.00
8) Nitrobenzene-d5	8.883	82	23978	0.350	ng	-0.02
11) 2-Methylnaphthalene-d10	12.106	152	53629	0.345	ng	-0.01
14) 2,4,6-Tribromophenol	15.855	330	6899	0.307	ng	-0.01
15) 2-Fluorobiphenyl	12.988	172	81190	0.446	ng	-0.01
27) Fluoranthene-d10	19.143	212	104573	0.349	ng	-0.01
31) Terphenyl-d14	19.753	244	121944	0.594	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.361	88	38040m	4.512	ng	Qvalue
25) Phenanthrene	17.152	178	5920	0.020	ng	98
28) Fluoranthene	19.173	202	7842	0.024	ng	# 94
34) Bis(2-ethylhexyl)phtha...	21.239	149	25542	0.117	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
Data File : BN014982.D
Acq On : 22 Jun 2021 20:16
Operator : CG/JU
Sample : M2750-11
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP01-20210615

Manual Integrations
APPROVED

mohammad
6/23/2021 4:50:31 PM

Quant Time: Jun 23 03:39:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 17 01:26:44 2021
Response via : Initial Calibration

