

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
 Data File : BN014995.D
 Acq On : 23 Jun 2021 06:57
 Operator : CG/JU
 Sample : M2750-22
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW179D2-20210616

Manual Integrations
 APPROVED

mohammad
 6/23/2021 4:50:52 PM

Quant Time: Jun 23 09:10:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 01:26:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	21442	0.400	ng	#-0.01
7) Naphthalene-d8	10.518	136	96522	0.400	ng	# 0.00
13) Acenaphthene-d10	14.359	164	51610	0.400	ng	-0.01
19) Phenanthrene-d10	17.104	188	99218	0.400	ng	-0.01
29) Chrysene-d12	21.303	240	88381	0.400	ng	#-0.02
35) Perylene-d12	23.551	264	77509	0.400	ng	-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.381	112	9952	0.150	ng	0.00
5) Phenol-d6	6.920	99	6949	0.083	ng	0.00
8) Nitrobenzene-d5	8.883	82	26902	0.369	ng	-0.02
11) 2-Methylnaphthalene-d10	12.102	152	55681	0.336	ng	-0.02
14) 2,4,6-Tribromophenol	15.843	330	7444	0.307	ng	-0.02
15) 2-Fluorobiphenyl	12.976	172	85048	0.434	ng	-0.02
27) Fluoranthene-d10	19.138	212	109975	0.352	ng	-0.02
31) Terphenyl-d14	19.744	244	133443	0.609	ng	-0.02
Target Compounds						Qvalue
2) 1,4-Dioxane	3.361	88	15356m	1.708	ng	
6) bis(2-Chloroethyl)ether	7.179	93	4268	0.065	ng	# 85
34) Bis(2-ethylhexyl)phtha...	21.229	149	26977	0.116	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
Data File : BN014995.D
Acq On : 23 Jun 2021 06:57
Operator : CG/JU
Sample : M2750-22
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW179D2-20210616

Manual Integrations
APPROVED

mohammad
6/23/2021 4:50:52 PM

Quant Time: Jun 23 09:10:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 17 01:26:44 2021
Response via : Initial Calibration

