

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
 Data File : BN014998.D
 Acq On : 23 Jun 2021 08:44
 Operator : CG/JU
 Sample : M2750-12
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE123D1-20210616

Manual Integrations
 APPROVED

mohammad
 6/23/2021 4:51:00 PM

Quant Time: Jun 23 09:44:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 17 01:26:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	19798	0.400	ng	#-0.01
7) Naphthalene-d8	10.505	136	89047	0.400	ng	#-0.02
13) Acenaphthene-d10	14.358	164	47077	0.400	ng	-0.01
19) Phenanthrene-d10	17.103	188	91609	0.400	ng	#-0.01
29) Chrysene-d12	21.302	240	79390	0.400	ng	#-0.02
35) Perylene-d12	23.554	264	68944	0.400	ng	-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.380	112	9560	0.156	ng	0.00
5) Phenol-d6	6.920	99	6499	0.084	ng	0.00
8) Nitrobenzene-d5	8.883	82	26098	0.388	ng	-0.02
11) 2-Methylnaphthalene-d10	12.101	152	53275	0.349	ng	-0.02
14) 2,4,6-Tribromophenol	15.843	330	7444	0.337	ng	-0.03
15) 2-Fluorobiphenyl	12.975	172	82008	0.459	ng	-0.03
27) Fluoranthene-d10	19.138	212	107225	0.371	ng	-0.02
31) Terphenyl-d14	19.744	244	120246	0.611	ng	-0.02
Target Compounds						Qvalue
2) 1,4-Dioxane	3.361	88	25035m	3.015	ng	
34) Bis(2-ethylhexyl)phtha...	21.228	149	35201	0.169	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062321\
Data File : BN014998.D
Acq On : 23 Jun 2021 08:44
Operator : CG/JU
Sample : M2750-12
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE123D1-20210616

Manual Integrations
APPROVED

mohammad
6/23/2021 4:51:00 PM

Quant Time: Jun 23 09:44:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 17 01:26:44 2021
Response via : Initial Calibration

