

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062420\
 Data File : BN011045.D
 Acq On : 24 Jun 2020 13:14
 Operator : CG/JU
 Sample : L3033-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 Z1-DUP-0571-200615

Quant Time: Jun 24 13:50:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 23 18:53:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	2640	0.40	ng	0.00
7) Naphthalene-d8	10.27	136	9249	0.40	ng	0.00
13) Acenaphthene-d10	14.14	164	5255	0.40	ng	0.00
19) Phenanthrene-d10	16.90	188	11076	0.40	ng	0.00
29) Chrysene-d12	21.11	240	10901	0.40	ng	-0.01
36) Perylene-d12	23.25	264	11175	0.40	ng	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.19	112	1353	0.18	ng	0.00
5) Phenol-d6	6.73	99	762	0.09	ng	0.02
8) Nitrobenzene-d5	8.66	82	2843	0.34	ng	0.01
11) 2-Methylnaphthalene-d10	11.86	152	4745	0.29	ng	0.00
14) 2,4,6-Tribromophenol	15.64	330	832	0.36	ng	0.00
15) 2-Fluorobiphenyl	12.76	172	9313	0.38	ng	0.00
27) Fluoranthene-d10	18.93	212	11797	0.34	ng	0.00
31) Terphenyl-d14	19.55	244	13549	0.46	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.23	88	7901	2.226	ng	99
6) bis(2-Chloroethyl)ether	6.97	93	989	0.130	ng	96
17) Acenaphthene	14.21	154	9665	0.543	ng	98
26) Anthracene	17.03	178	1322	0.043	ng	# 87
28) Fluoranthene	18.96	202	1284	0.030	ng	# 86
30) Pyrene	19.33	202	1538	0.037	ng	# 84
34) Bis(2-ethylhexyl)phthalate	21.06	149	7860	0.518	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062420\
Data File : BN011045.D
Acq On : 24 Jun 2020 13:14
Operator : CG/JU
Sample : L3033-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
Z1-DUP-0571-200615

Quant Time: Jun 24 13:50:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 23 18:53:50 2020
Response via : Initial Calibration

