

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
 Data File : BN015012.D
 Acq On : 23 Jun 2021 17:06
 Operator : CG/JU
 Sample : PB137284BSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB137284BSD

Quant Time: Jun 24 01:10:23 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 23 16:24:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	20728	0.40	ng	# 0.00
7) Naphthalene-d8	10.518	136	90115	0.40	ng	# 0.01
13) Acenaphthene-d10	14.358	164	46516	0.40	ng	0.00
19) Phenanthrene-d10	17.104	188	88087	0.40	ng	# 0.00
29) Chrysene-d12	21.303	240	79577	0.40	ng	# 0.00
35) Perylene-d12	23.550	264	71250	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.371	112	22229	0.35	ng	0.00
5) Phenol-d6	6.911	99	26842	0.33	ng	0.00
8) Nitrobenzene-d5	8.883	82	21666	0.32	ng	0.00
11) 2-Methylnaphthalene-d10	12.101	152	54473	0.35	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	6030	0.28	ng	0.00
15) 2-Fluorobiphenyl	12.975	172	70665	0.40	ng	0.00
27) Fluoranthene-d10	19.138	212	93936	0.34	ng	0.00
31) Terphenyl-d14	19.744	244	72009	0.37	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.352	88	3596	0.41	ng	# 71
3) n-Nitrosodimethylamine	3.702	42	6508	0.34	ng	# 72
6) bis(2-Chloroethyl)ether	7.178	93	24761	0.39	ng	# 84
9) Naphthalene	10.556	128	96682	0.38	ng	97
10) Hexachlorobutadiene	10.860	225	18732	0.37	ng	# 99
12) 2-Methylnaphthalene	12.173	142	62887	0.37	ng	# 89
16) Acenaphthylene	14.079	152	76091	0.32	ng	98
17) Acenaphthene	14.422	154	53560	0.37	ng	98
18) Fluorene	15.411	166	64810	0.37	ng	99
20) 4,6-Dinitro-2-methylph...	15.475	198	1079	0.30	ng	# 57
21) 4-Bromophenyl-phenylether	16.300	248	20233	0.36	ng	89
22) Hexachlorobenzene	16.410	284	23339	0.39	ng	# 93
23) Atrazine	16.568	200	12622	0.25	ng	94
24) Pentachlorophenol	16.751	266	5272	0.24	ng	99
25) Phenanthrene	17.140	178	105654	0.39	ng	99
26) Anthracene	17.237	178	89744	0.34	ng	99
28) Fluoranthene	19.168	202	115685	0.38	ng	# 97
30) Pyrene	19.530	202	114821	0.36	ng	99
32) Benzo(a)anthracene	21.281	228	100867	0.34	ng	98
33) Chrysene	21.334	228	110386	0.38	ng	98
34) Bis(2-ethylhexyl)phtha...	21.228	149	33538	0.16	ng	# 87
36) Indeno(1,2,3-cd)pyrene	25.813	276	83631	0.35	ng	# 86
37) Benzo(b)fluoranthene	22.876	252	99370	0.37	ng	# 96
38) Benzo(k)fluoranthene	22.921	252	109250	0.42	ng	# 94
39) Benzo(a)pyrene	23.454	252	84819	0.36	ng	# 94
40) Dibenzo(a,h)anthracene	25.830	278	64879	0.34	ng	# 91
41) Benzo(g,h,i)perylene	26.501	276	68657	0.33	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
Data File : BN015012.D
Acq On : 23 Jun 2021 17:06
Operator : CG/JU
Sample : PB137284BSD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB137284BSD

Quant Time: Jun 24 01:10:23 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 16:24:59 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.750 min Scan# 485
Delta R.T. 0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion: 152 Resp: 20728
Ion Ratio Lower Upper
152 100
150 155.5 119.0 178.6
115 58.9 62.2 93.4#



#2
1,4-Dioxane
Concen: 0.41 ng
RT: 3.352 min Scan# 8
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion: 88 Resp: 3596
Ion Ratio Lower Upper
88 100
43 24.8 12.9 19.3#
58 68.6 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.34 ng
RT: 3.702 min Scan# 46
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion: 42 Resp: 6508
Ion Ratio Lower Upper
42 100
74 145.6 149.5 224.3#
44 6.7 4.2 6.2#



#4
2-Fluorophenol
Concen: 0.35 ng
RT: 5.371 min Scan# 227
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion: 112 Resp: 22229
Ion Ratio Lower Upper
112 100
64 57.2 39.0 58.6
63 28.5 20.6 30.8









#5
Phenol-d6
Concen: 0.33 ng
RT: 6.911 min Scan# 394
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion: 99 Resp: 26842
Ion Ratio Lower Upper
99 100
42 17.7 8.9 13.3#
71 30.8 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.178 min Scan# 423
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion: 93 Resp: 24761
Ion Ratio Lower Upper
93 100
63 75.8 46.6 70.0#
95 32.9 28.6 42.8









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.518 min Scan# 720
Delta R.T. 0.012 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:	136	Resp:	90115
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.8	14.8
54	4.6	5.4	8.2#
68	4.0	4.9	7.3#



#8
Nitrobenzene-d5
Concen: 0.32 ng
RT: 8.883 min Scan# 591
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:	82	Resp:	21666
Ion Ratio	Lower	Upper	
82	100		
128	40.8	31.6	47.4
54	47.7	29.0	43.4#









#9
Naphthalene
Concen: 0.38 ng
RT: 10.556 min Scan# 723
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:	128	Resp:	96682
Ion Ratio	Lower	Upper	
128	100		
129	11.0	9.1	13.7
127	13.2	11.8	17.6



#10
Hexachlorobutadiene
Concen: 0.37 ng
RT: 10.860 min Scan# 747
Delta R.T. 0.012 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:	225	Resp:	18732
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.8	50.2	75.2









#11
2-Methylnaphthalene-d10
Concen: 0.35 ng
RT: 12.101 min Scan# 930
Delta R.T. 0.004 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:152 Resp: 54473
Ion Ratio Lower Upper
152 100
151 20.5 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.173 min Scan# 946
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:142 Resp: 62887
Ion Ratio Lower Upper
142 100
141 88.1 72.3 108.5
115 33.3 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.358 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:164 Resp: 46516
Ion Ratio Lower Upper
164 100
162 100.4 83.8 125.6
160 44.9 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.28 ng
RT: 15.843 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:330 Resp: 6030
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.0 22.4 33.6#









#15
2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.975 min Scan# 1058
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:172	Resp:	70665
Ion Ratio	Lower	Upper
172	100	
171	35.3	30.6 45.8
170	23.7	20.2 30.4



#16
Acenaphthylene
Concen: 0.32 ng
RT: 14.079 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:152	Resp:	76091
Ion Ratio	Lower	Upper
152	100	
151	19.4	16.7 25.1
153	13.1	10.5 15.7









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.422 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:154 Resp: 53560
Ion Ratio Lower Upper
154 100
153 112.5 90.9 136.3
152 54.8 45.8 68.8



#18
Fluorene
Concen: 0.37 ng
RT: 15.411 min Scan# 1250
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:166 Resp: 64810
Ion Ratio Lower Upper
166 100
165 97.3 79.2 118.8
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.104 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:188 Resp: 88087
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.2 9.0 13.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.30 ng
RT: 15.475 min Scan# 1255
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:198 Resp: 1079
Ion Ratio Lower Upper
198 100
182 15.7 34.5 51.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.300 min Scan# 1321
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:248 Resp: 20233
Ion Ratio Lower Upper
248 100
250 97.3 86.2 129.4
141 53.4 36.2 54.2



#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.410 min Scan# 1330
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:284 Resp: 23339
Ion Ratio Lower Upper
284 100
142 33.7 22.2 33.2#
249 31.0 26.6 39.8









#23
Atrazine
Concen: 0.25 ng
RT: 16.568 min Scan# 1343
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:200 Resp: 12622
Ion Ratio Lower Upper
200 100
173 26.2 21.7 32.5
215 48.0 43.1 64.7



#24
Pentachlorophenol
Concen: 0.24 ng
RT: 16.751 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:266 Resp: 5272
Ion Ratio Lower Upper
266 100
264 62.6 51.0 76.4
268 64.0 50.6 76.0









#25
Phenanthrene
Concen: 0.39 ng
RT: 17.140 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:178 Resp: 105654
Ion Ratio Lower Upper
178 100
176 18.7 15.6 23.4
179 15.5 12.1 18.1



#26
Anthracene
Concen: 0.34 ng
RT: 17.237 min Scan# 1398
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:178 Resp: 89744
Ion Ratio Lower Upper
178 100
176 18.2 15.1 22.7
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.34 ng
RT: 19.138 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:212 Resp: 93936
Ion Ratio Lower Upper
212 100
106 14.1 7.0 10.6#
104 7.8 4.1 6.1#



#28
Fluoranthene
Concen: 0.38 ng
RT: 19.168 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:202 Resp: 115685
Ion Ratio Lower Upper
202 100
101 11.7 6.6 10.0#
203 17.4 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.303 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:240 Resp: 79577
Ion Ratio Lower Upper
240 100
120 7.7 4.2 6.4#
236 25.8 21.4 32.2



#30
Pyrene
Concen: 0.36 ng
RT: 19.530 min Scan# 1767
Delta R.T. -0.005 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:202 Resp: 114821
Ion Ratio Lower Upper
202 100
200 20.4 16.9 25.3
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.37 ng
RT: 19.744 min Scan# 1810
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:244 Resp: 72009
Ion Ratio Lower Upper
244 100
212 7.6 6.3 9.5
122 13.1 6.2 9.2#



#32
Benzo(a)anthracene
Concen: 0.34 ng
RT: 21.281 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:228 Resp: 100867
Ion Ratio Lower Upper
228 100
226 27.3 23.0 34.6
229 19.2 15.7 23.5









#33
Chrysene
Concen: 0.38 ng
RT: 21.334 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:228 Resp: 110386
Ion Ratio Lower Upper
228 100
226 29.7 25.0 37.6
229 19.2 15.9 23.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.16 ng
RT: 21.228 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:149 Resp: 33538
Ion Ratio Lower Upper
149 100
167 28.1 28.9 43.3#
279 3.6 4.9 7.3#









#35
Perylene-d12
Concen: 0.40 ng
RT: 23.550 min Scan# 2423
Delta R.T. -0.004 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:264 Resp: 71250
Ion Ratio Lower Upper
264 100
260 24.7 20.6 31.0
265 24.2 21.4 32.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.35 ng
RT: 25.813 min Scan# 3084
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:276 Resp: 83631
Ion Ratio Lower Upper
276 100
138 28.5 10.0 15.0#
277 25.0 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.37 ng
RT: 22.876 min Scan# 2226
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:252 Resp: 99370
Ion Ratio Lower Upper
252 100
253 22.1 18.5 27.7
125 11.1 5.8 8.6#



#38
Benzo(k)fluoranthene
Concen: 0.42 ng
RT: 22.921 min Scan# 2239
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Tgt Ion:252 Resp: 109250
Ion Ratio Lower Upper
252 100
253 22.3 19.1 28.7
125 12.0 5.4 8.0#









#39
Benzo(a)pyrene
Concen: 0.36 ng
RT: 23.454 min Scan# 2395
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:252 Resp: 84819
Ion Ratio Lower Upper
252 100
253 22.2 19.4 29.2
125 12.8 7.4 11.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.34 ng
RT: 25.830 min Scan# 3089
Delta R.T. -0.000 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06


Tgt Ion:278 Resp: 64879
Ion Ratio Lower Upper
278 100
139 18.6 9.0 13.4#
279 24.0 21.0 31.4





#41
Benzo(g,h,i)perylene
Concen: 0.33 ng
RT: 26.501 min Scan# 3285
Delta R.T. -0.004 min
Lab File: BN015012.D
Acq: 23 Jun 2021 17:06

Instrument :
BNA_N
ClientSampleId :
PB137284BSD



Tgt Ion: 276 Resp: 68657
Ion Ratio Lower Upper
276 100
277 24.0 19.7 29.5
138 23.4 9.8 14.6#

