

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
 Data File : BN015026.D
 Acq On : 24 Jun 2021 01:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Jun 24 03:21:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 23 16:24:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	22221	0.40	ng	# 0.00
7) Naphthalene-d8	10.505	136	97769	0.40	ng	# 0.00
13) Acenaphthene-d10	14.358	164	51891	0.40	ng	0.00
19) Phenanthrene-d10	17.103	188	98479	0.40	ng	# 0.00
29) Chrysene-d12	21.292	240	91415	0.40	ng	#-0.01
35) Perylene-d12	23.543	264	84499	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.371	112	25503	0.37	ng	0.00
5) Phenol-d6	6.911	99	31097	0.36	ng	0.00
8) Nitrobenzene-d5	8.883	82	27054	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.097	152	59923	0.36	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	8577	0.35	ng	0.00
15) 2-Fluorobiphenyl	12.975	172	77201	0.39	ng	0.00
27) Fluoranthene-d10	19.138	212	108637	0.35	ng	0.00
31) Terphenyl-d14	19.739	244	81531	0.36	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.352	88	3997	0.43	ng	# 73
3) n-Nitrosodimethylamine	3.702	42	7309	0.35	ng	75
6) bis(2-Chloroethyl)ether	7.178	93	26873	0.39	ng	# 84
9) Naphthalene	10.556	128	104804	0.38	ng	97
10) Hexachlorobutadiene	10.847	225	20351	0.37	ng	# 99
12) 2-Methylnaphthalene	12.172	142	68809	0.37	ng	# 89
16) Acenaphthylene	14.066	152	89530	0.34	ng	98
17) Acenaphthene	14.422	154	59699	0.37	ng	98
18) Fluorene	15.411	166	72867	0.37	ng	99
20) 4,6-Dinitro-2-methylph...	15.475	198	633	0.21	ng	# 67
21) 4-Bromophenyl-phenylether	16.300	248	22970	0.36	ng	93
22) Hexachlorobenzene	16.410	284	25713	0.38	ng	# 93
23) Atrazine	16.568	200	16716	0.29	ng	93
24) Pentachlorophenol	16.751	266	8209	0.33	ng	99
25) Phenanthrene	17.140	178	115762	0.39	ng	99
26) Anthracene	17.237	178	101718	0.35	ng	99
28) Fluoranthene	19.168	202	131574	0.39	ng	# 97
30) Pyrene	19.530	202	133153	0.37	ng	99
32) Benzo(a)anthracene	21.281	228	121134	0.35	ng	97
33) Chrysene	21.334	228	126776	0.38	ng	97
34) Bis(2-ethylhexyl)phtha...	21.228	149	60451	0.25	ng	# 87
36) Indeno(1,2,3-cd)pyrene	25.799	276	104155	0.37	ng	# 85
37) Benzo(b)fluoranthene	22.869	252	123188	0.39	ng	# 95
38) Benzo(k)fluoranthene	22.910	252	127193	0.42	ng	# 94
39) Benzo(a)pyrene	23.444	252	106127	0.38	ng	# 92
40) Dibenzo(a,h)anthracene	25.816	278	82444	0.36	ng	# 91
41) Benzo(g,h,i)perylene	26.487	276	84234	0.34	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
Data File : BN015026.D
Acq On : 24 Jun 2021 01:28
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Jun 24 03:21:56 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 16:24:59 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.750 min Scan# 485
 Delta R.T. -0.000 min
 Lab File: BN015026.D
 Acq: 24 Jun 2021 01:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion: 152 Resp: 22221
 Ion Ratio Lower Upper
 152 100
 150 154.1 119.0 178.6
 115 58.7 62.2 93.4#



#2
 1,4-Dioxane
 Concen: 0.43 ng
 RT: 3.352 min Scan# 8
 Delta R.T. -0.000 min
 Lab File: BN015026.D
 Acq: 24 Jun 2021 01:28

Tgt Ion: 88 Resp: 3997
 Ion Ratio Lower Upper
 88 100
 43 23.9 12.9 19.3#
 58 67.4 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.35 ng
RT: 3.702 min Scan# 46
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	42	Resp:	7309
Ion	Ratio	Lower	Upper
42	100		
74	150.0	149.5	224.3
44	6.0	4.2	6.2



#4
2-Fluorophenol
Concen: 0.37 ng
RT: 5.371 min Scan# 227
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:	112	Resp:	25503
Ion	Ratio	Lower	Upper
112	100		
64	57.2	39.0	58.6
63	28.5	20.6	30.8









#5
Phenol-d6
Concen: 0.36 ng
RT: 6.911 min Scan# 394
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 31097
Ion Ratio Lower Upper
99 100
42 17.2 8.9 13.3#
71 30.9 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 0.39 ng
RT: 7.178 min Scan# 423
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion: 93 Resp: 26873
Ion Ratio Lower Upper
93 100
63 75.5 46.6 70.0#
95 32.9 28.6 42.8









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.505 min Scan# 719
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	97769
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.8	14.8
54	8.7	5.4	8.2#
68	7.1	4.9	7.3



#8
Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.883 min Scan# 591
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:	82	Resp:	27054
Ion Ratio	Lower	Upper	
82	100		
128	44.8	31.6	47.4
54	46.4	29.0	43.4#









#9
Naphthalene
Concen: 0.38 ng
RT: 10.556 min Scan# 723
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 104804
Ion Ratio Lower Upper
128 100
129 11.0 9.1 13.7
127 13.0 11.8 17.6



#10
Hexachlorobutadiene
Concen: 0.37 ng
RT: 10.847 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:225 Resp: 20351
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 50.2 75.2









#11
2-Methylnaphthalene-d10
Concen: 0.36 ng
RT: 12.097 min Scan# 929
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 59923
Ion Ratio Lower Upper
152 100
151 20.3 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.37 ng
RT: 12.172 min Scan# 946
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:142 Resp: 68809
Ion Ratio Lower Upper
142 100
141 87.7 72.3 108.5
115 32.6 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.358 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 51891
Ion Ratio Lower Upper
164 100
162 99.0 83.8 125.6
160 43.5 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.35 ng
RT: 15.843 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:330 Resp: 8577
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.8 22.4 33.6#









#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.975 min Scan# 1058
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172	Resp:	77201
Ion Ratio	Lower	Upper
172	100	
171	35.1	30.6 45.8
170	23.5	20.2 30.4



#16
Acenaphthylene
Concen: 0.34 ng
RT: 14.066 min Scan# 1144
Delta R.T. -0.013 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:152	Resp:	89530
Ion Ratio	Lower	Upper
152	100	
151	19.3	16.7 25.1
153	13.0	10.5 15.7









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.422 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 59699
Ion Ratio Lower Upper
154 100
153 112.4 90.9 136.3
152 54.2 45.8 68.8



#18
Fluorene
Concen: 0.37 ng
RT: 15.411 min Scan# 1250
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:166 Resp: 72867
Ion Ratio Lower Upper
166 100
165 97.5 79.2 118.8
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.103 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 98479
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.3 9.0 13.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.21 ng
RT: 15.475 min Scan# 1255
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:198 Resp: 633
Ion Ratio Lower Upper
198 100
182 22.2 34.5 51.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.300 min Scan# 1321
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 22970
Ion Ratio Lower Upper
248 100
250 98.9 86.2 129.4
141 47.3 36.2 54.2



#22
Hexachlorobenzene
Concen: 0.38 ng
RT: 16.410 min Scan# 1330
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:284 Resp: 25713
Ion Ratio Lower Upper
284 100
142 33.5 22.2 33.2#
249 30.8 26.6 39.8









#23
Atrazine
Concen: 0.29 ng
RT: 16.568 min Scan# 1343
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 16716
Ion Ratio Lower Upper
200 100
173 24.8 21.7 32.5
215 48.4 43.1 64.7



#24
Pentachlorophenol
Concen: 0.33 ng
RT: 16.751 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:266 Resp: 8209
Ion Ratio Lower Upper
266 100
264 62.7 51.0 76.4
268 64.2 50.6 76.0









#25
Phenanthrene
Concen: 0.39 ng
RT: 17.140 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 115762
Ion Ratio Lower Upper
178 100
176 18.8 15.6 23.4
179 15.5 12.1 18.1



#26
Anthracene
Concen: 0.35 ng
RT: 17.237 min Scan# 1398
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:178 Resp: 101718
Ion Ratio Lower Upper
178 100
176 18.2 15.1 22.7
179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.35 ng
RT: 19.138 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 108637
Ion Ratio Lower Upper
212 100
106 14.2 7.0 10.6#
104 7.8 4.1 6.1#



#28
Fluoranthene
Concen: 0.39 ng
RT: 19.168 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:202 Resp: 131574
Ion Ratio Lower Upper
202 100
101 11.7 6.6 10.0#
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.292 min Scan# 1984
Delta R.T. -0.011 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 91415
Ion Ratio Lower Upper
240 100
120 12.6 4.2 6.4#
236 27.2 21.4 32.2



#30
Pyrene
Concen: 0.37 ng
RT: 19.530 min Scan# 1767
Delta R.T. -0.005 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:202 Resp: 133153
Ion Ratio Lower Upper
202 100
200 20.5 16.9 25.3
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.36 ng
RT: 19.739 min Scan# 1809
Delta R.T. -0.005 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 81531
Ion Ratio Lower Upper
244 100
212 7.8 6.3 9.5
122 14.3 6.2 9.2#



#32
Benzo(a)anthracene
Concen: 0.35 ng
RT: 21.281 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:228 Resp: 121134
Ion Ratio Lower Upper
228 100
226 25.8 23.0 34.6
229 19.7 15.7 23.5









#33
Chrysene
Concen: 0.38 ng
RT: 21.334 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	228	Resp:	126776
Ion Ratio	Lower	Upper	
228	100		
226	28.4	25.0	37.6
229	20.0	15.9	23.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.25 ng
RT: 21.228 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:	149	Resp:	60451
Ion Ratio	Lower	Upper	
149	100		
167	28.2	28.9	43.3#
279	3.7	4.9	7.3#









#35
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.543 min Scan# 2421
 Delta R.T. -0.011 min
 Lab File: BN015026.D
 Acq: 24 Jun 2021 01:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:264 Resp: 84499
 Ion Ratio Lower Upper
 264 100
 260 24.7 20.6 31.0
 265 23.6 21.4 32.2



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.37 ng
 RT: 25.799 min Scan# 3080
 Delta R.T. -0.014 min
 Lab File: BN015026.D
 Acq: 24 Jun 2021 01:28

Tgt Ion:276 Resp: 104155
 Ion Ratio Lower Upper
 276 100
 138 30.3 10.0 15.0#
 277 25.0 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.39 ng
RT: 22.869 min Scan# 2224
Delta R.T. -0.007 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 123188
Ion Ratio Lower Upper
252 100
253 22.0 18.5 27.7
125 12.0 5.8 8.6#



#38
Benzo(k)fluoranthene
Concen: 0.42 ng
RT: 22.910 min Scan# 2236
Delta R.T. -0.011 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:252 Resp: 127193
Ion Ratio Lower Upper
252 100
253 22.2 19.1 28.7
125 12.1 5.4 8.0#









#39
Benzo(a)pyrene
Concen: 0.38 ng
RT: 23.444 min Scan# 2392
Delta R.T. -0.011 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 106127
Ion Ratio Lower Upper
252 100
253 22.6 19.4 29.2
125 17.1 7.4 11.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.36 ng
RT: 25.816 min Scan# 3085
Delta R.T. -0.014 min
Lab File: BN015026.D
Acq: 24 Jun 2021 01:28

Tgt Ion:278 Resp: 82444
Ion Ratio Lower Upper
278 100
139 18.6 9.0 13.4#
279 23.9 21.0 31.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.34 ng
 RT: 26.487 min Scan# 3281
 Delta R.T. -0.017 min
 Lab File: BN015026.D
 Acq: 24 Jun 2021 01:28

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:276 Resp: 84234
 Ion Ratio Lower Upper
 276 100
 277 23.5 19.7 29.5
 138 23.8 9.8 14.6#

