

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
 Data File : BN015016.D
 Acq On : 23 Jun 2021 19:30
 Operator : CG/JU
 Sample : M2799-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SB-1-GW

Quant Time: Jun 24 01:11:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 23 16:24:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	21288	0.400	ng	# 0.00
7) Naphthalene-d8	10.505	136	93861	0.400	ng	# 0.00
13) Acenaphthene-d10	14.359	164	56190	0.400	ng	0.00
19) Phenanthrene-d10	17.104	188	103694	0.400	ng	# 0.00
29) Chrysene-d12	21.303	240	95177	0.400	ng	# 0.00
35) Perylene-d12	23.551	264	91689	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.381	112	9907	0.150	ng	0.00
5) Phenol-d6	6.920	99	8674	0.104	ng	0.00
8) Nitrobenzene-d5	8.883	82	28991	0.409	ng	0.00
11) 2-Methylnaphthalene-d10	12.097	152	53903	0.335	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	10033	0.380	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	80515	0.377	ng	0.00
27) Fluoranthene-d10	19.138	212	109325	0.334	ng	0.00
31) Terphenyl-d14	19.744	244	98367	0.417	ng	0.00
Target Compounds						
						Qvalue
9) Naphthalene	10.556	128	940533	3.575	ng	97
12) 2-Methylnaphthalene	12.173	142	158104	0.893	ng	# 89
16) Acenaphthylene	14.067	152	58857	0.207	ng	# 97
17) Acenaphthene	14.422	154	4624	0.027	ng	# 88
18) Fluorene	15.412	166	10370	0.048	ng	94
25) Phenanthrene	17.140	178	17554	0.056	ng	96
26) Anthracene	17.238	178	12745	0.041	ng	# 94
28) Fluoranthene	19.168	202	12747	0.036	ng	# 50
30) Pyrene	19.531	202	10292	0.027	ng	# 94
32) Benzo(a)anthracene	21.282	228	7244	0.020	ng	# 74
34) Bis(2-ethylhexyl)phtha...	21.228	149	56553	0.226	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
Data File : BN015016.D
Acq On : 23 Jun 2021 19:30
Operator : CG/JU
Sample : M2799-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SB-1-GW

Quant Time: Jun 24 01:11:09 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 16:24:59 2021
Response via : Initial Calibration

