

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
 Data File : BN015017.D
 Acq On : 23 Jun 2021 20:06
 Operator : CG/JU
 Sample : M2799-03
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SB-2-GW

Quant Time: Jun 24 01:11:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 23 16:24:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	20878	0.400	ng	# 0.00
7) Naphthalene-d8	10.505	136	94543	0.400	ng	# 0.00
13) Acenaphthene-d10	14.358	164	49852	0.400	ng	0.00
19) Phenanthrene-d10	17.104	188	97123	0.400	ng	# 0.00
29) Chrysene-d12	21.303	240	85319	0.400	ng	# 0.00
35) Perylene-d12	23.547	264	74295	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.380	112	2901	0.045	ng	0.00
5) Phenol-d6	6.920	99	3610	0.044	ng	0.00
8) Nitrobenzene-d5	8.883	82	12324	0.172	ng	0.00
11) 2-Methylnaphthalene-d10	12.097	152	23655	0.146	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	1087	0.046	ng	0.00
15) 2-Fluorobiphenyl	12.975	172	33035	0.174	ng	0.00
27) Fluoranthene-d10	19.138	212	27674	0.090	ng	0.00
31) Terphenyl-d14	19.744	244	20326	0.096	ng	0.00
Target Compounds						
9) Naphthalene	10.556	128	390767	1.475	ng	Qvalue 97
12) 2-Methylnaphthalene	12.173	142	36291	0.204	ng	# 89
16) Acenaphthylene	14.067	152	19733	0.078	ng	98
34) Bis(2-ethylhexyl)phtha...	21.228	149	6128	0.027	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
Data File : BN015017.D
Acq On : 23 Jun 2021 20:06
Operator : CG/JU
Sample : M2799-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SB-2-GW

Quant Time: Jun 24 01:11:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 16:24:59 2021
Response via : Initial Calibration

