

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
 Data File : BN015035.D
 Acq On : 24 Jun 2021 10:19
 Operator : CG/JU
 Sample : M2784-02DL 2X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE137-20210618DL

Manual Integrations
 APPROVED

mohammad
 6/24/2021 2:57:32 PM

Quant Time: Jun 24 10:49:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 23 16:24:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	21118	0.400	ng	#-0.02
7) Naphthalene-d8	10.505	136	93993	0.400	ng	# 0.00
13) Acenaphthene-d10	14.346	164	49381	0.400	ng	-0.01
19) Phenanthrene-d10	17.091	188	95573	0.400	ng	#-0.01
29) Chrysene-d12	21.303	240	80411	0.400	ng	# 0.00
35) Perylene-d12	23.550	264	67800	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.362	112	5750	0.088	ng	0.00
5) Phenol-d6	6.911	99	4099	0.050	ng	0.00
8) Nitrobenzene-d5	8.870	82	13097	0.184	ng	-0.01
11) 2-Methylnaphthalene-d10	12.093	152	28899	0.179	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	3235	0.140	ng	0.00
15) 2-Fluorobiphenyl	12.975	172	43001	0.229	ng	0.00
27) Fluoranthene-d10	19.138	212	55076	0.183	ng	0.00
31) Terphenyl-d14	19.744	244	49156	0.247	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.343	88	22636m	2.556	ng	
6) bis(2-Chloroethyl)ether	7.169	93	1855	0.029	ng	# 86
34) Bis(2-ethylhexyl)phtha...	21.228	149	22596	0.107	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
Data File : BN015035.D
Acq On : 24 Jun 2021 10:19
Operator : CG/JU
Sample : M2784-02DL 2X
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE137-20210618DL

Manual Integrations
APPROVED

mohammad
6/24/2021 2:57:32 PM

Quant Time: Jun 24 10:49:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 16:24:59 2021
Response via : Initial Calibration

