

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
 Data File : BN015028.D
 Acq On : 24 Jun 2021 04:25
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad
 6/24/2021 9:53:25 AM

Quant Time: Jun 24 05:43:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 23 16:24:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	23790	0.400	ng	# 0.00
7) Naphthalene-d8	10.505	136	104884	0.400	ng	0.00
13) Acenaphthene-d10	14.358	164	55711	0.400	ng	0.00
19) Phenanthrene-d10	17.103	188	107486	0.400	ng	# 0.00
29) Chrysene-d12	21.292	240	96656	0.400	ng	#-0.01
35) Perylene-d12	23.540	264	87145	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.371	112	26808	0.364	ng	0.00
5) Phenol-d6	6.911	99	32539	0.349	ng	0.00
8) Nitrobenzene-d5	8.883	82	27172	0.343	ng	0.00
11) 2-Methylnaphthalene-d10	12.097	152	63847	0.355	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	8338	0.319	ng	0.00
15) 2-Fluorobiphenyl	12.975	172	83114	0.393	ng	0.00
27) Fluoranthene-d10	19.138	212	114483	0.338	ng	0.00
31) Terphenyl-d14	19.739	244	87283	0.364	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.352	88	3624m	0.363	ng	
3) n-Nitrosodimethylamine	3.702	42	7787	0.352	ng	# 73
6) bis(2-Chloroethyl)ether	7.178	93	28852	0.394	ng	# 84
9) Naphthalene	10.556	128	111918	0.381	ng	97
10) Hexachlorobutadiene	10.847	225	21700	0.372	ng	# 98
12) 2-Methylnaphthalene	12.172	142	73645	0.372	ng	# 88
16) Acenaphthylene	14.066	152	92700	0.329	ng	98
17) Acenaphthene	14.422	154	63614	0.369	ng	98
18) Fluorene	15.399	166	77852	0.366	ng	99
20) 4,6-Dinitro-2-methylph...	15.475	198	741	0.219	ng	# 61
21) 4-Bromophenyl-phenylether	16.300	248	24794	0.357	ng	94
22) Hexachlorobenzene	16.410	284	27780	0.379	ng	# 93
23) Atrazine	16.568	200	16582	0.267	ng	94
24) Pentachlorophenol	16.750	266	7528	0.279	ng	99
25) Phenanthrene	17.140	178	125592	0.383	ng	99
26) Anthracene	17.225	178	110041	0.343	ng	99
28) Fluoranthene	19.163	202	139904	0.377	ng	# 97
30) Pyrene	19.525	202	141494	0.370	ng	99
32) Benzo(a)anthracene	21.281	228	126601	0.347	ng	97
33) Chrysene	21.334	228	133899	0.384	ng	97
34) Bis(2-ethylhexyl)phtha...	21.228	149	54889	0.216	ng	# 87
36) Indeno(1,2,3-cd)pyrene	25.799	276	102444	0.352	ng	# 84
37) Benzo(b)fluoranthene	22.869	252	127937	0.391	ng	# 95
38) Benzo(k)fluoranthene	22.910	252	132525	0.420	ng	# 94
39) Benzo(a)pyrene	23.444	252	110132	0.383	ng	# 91
40) Dibenzo(a,h)anthracene	25.816	278	79781	0.339	ng	# 91
41) Benzo(g,h,i)perylene	26.487	276	84188	0.333	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062421\
Data File : BN015028.D
Acq On : 24 Jun 2021 04:25
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM

Quant Time: Jun 24 05:43:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN061621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 16:24:59 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.750 min Scan# 485
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25





Tgt Ion: 152 Resp: 23790
Ion Ratio Lower Upper
152 100
150 155.9 119.0 178.6
115 59.3 62.2 93.4#



#2
1,4-Dioxane
Concen: 0.363 ng m
RT: 3.352 min Scan# 8
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Tgt Ion: 88 Resp: 3624
Ion Ratio Lower Upper
88 100
43 27.0 12.9 19.3#
58 77.2 38.1 57.1#



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/24/2021 9:53:25 AM







#3
n-Nitrosodimethylamine
Concen: 0.352 ng
RT: 3.702 min Scan# 46
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 42 Resp: 7787
Ion Ratio Lower Upper
42 100
74 147.1 149.5 224.3#
44 7.4 4.2 6.2#




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM



#4
2-Fluorophenol
Concen: 0.364 ng
RT: 5.371 min Scan# 227
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion: 112 Resp: 26808
Ion Ratio Lower Upper
112 100
64 57.2 39.0 58.6
63 28.6 20.6 30.8









#5
Phenol-d6
Concen: 0.349 ng
RT: 6.911 min Scan# 394
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 99 Resp: 32539
Ion Ratio Lower Upper
99 100
42 17.4 8.9 13.3#
71 31.0 35.6 53.4#




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.394 ng
RT: 7.178 min Scan# 423
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25


Tgt Ion: 93 Resp: 28852
Ion Ratio Lower Upper
93 100
63 75.8 46.6 70.0#
95 32.9 28.6 42.8





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.505 min Scan# 719
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:136 Resp: 104884
Ion Ratio Lower Upper
136 100
137 10.9 9.8 14.8
54 8.1 5.4 8.2
68 6.5 4.9 7.3

Manual Integrations
APPROVED



mohammad
6/24/2021 9:53:25 AM




#8
Nitrobenzene-d5
Concen: 0.343 ng
RT: 8.883 min Scan# 591
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25



Tgt Ion: 82 Resp: 27172
Ion Ratio Lower Upper
82 100
128 44.5 31.6 47.4
54 45.2 29.0 43.4#









#9
Naphthalene
Concen: 0.381 ng
RT: 10.556 min Scan# 723
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:128 Resp: 111918
Ion Ratio Lower Upper
128 100
129 11.0 9.1 13.7
127 13.0 11.8 17.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM



#10
Hexachlorobutadiene
Concen: 0.372 ng
RT: 10.847 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:225 Resp: 21700
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 50.2 75.2









#11
2-Methylnaphthalene-d10
Concen: 0.355 ng
RT: 12.097 min Scan# 929
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:152 Resp: 63847
Ion Ratio Lower Upper
152 100
151 20.3 16.7 25.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM



#12
2-Methylnaphthalene
Concen: 0.372 ng
RT: 12.172 min Scan# 946
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25


Tgt Ion:142 Resp: 73645
Ion Ratio Lower Upper
142 100
141 87.2 72.3 108.5
115 32.4 40.7 61.1#





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.358 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:164 Resp: 55711
Ion Ratio Lower Upper
164 100
162 99.0 83.8 125.6
160 43.3 37.4 56.0

Manual Integrations
APPROVED



mohammad
6/24/2021 9:53:25 AM




#14
2,4,6-Tribromophenol
Concen: 0.319 ng
RT: 15.843 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25



Tgt Ion:330 Resp: 8338
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.3 22.4 33.6#









#15
2-Fluorobiphenyl
Concen: 0.393 ng
RT: 12.975 min Scan# 1058
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:	172	Resp:	83114
Ion	Ratio	Lower	Upper
172	100		
171	34.6	30.6	45.8
170	23.3	20.2	30.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM



#16
Acenaphthylene
Concen: 0.329 ng
RT: 14.066 min Scan# 1144
Delta R.T. -0.013 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:	152	Resp:	92700
Ion	Ratio	Lower	Upper
152	100		
151	19.3	16.7	25.1
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.369 ng
RT: 14.422 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:154 Resp: 63614
Ion Ratio Lower Upper
154 100
153 112.7 90.9 136.3
152 54.8 45.8 68.8

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM



#18
Fluorene
Concen: 0.366 ng
RT: 15.399 min Scan# 1249
Delta R.T. -0.013 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25


Tgt Ion:166 Resp: 77852
Ion Ratio Lower Upper
166 100
165 97.7 79.2 118.8
167 13.5 10.7 16.1





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.103 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:188 Resp: 107486
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.0 9.0 13.4#

Manual Integrations
APPROVED



mohammad
6/24/2021 9:53:25 AM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.219 ng
RT: 15.475 min Scan# 1255
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Tgt Ion:198 Resp: 741
Ion Ratio Lower Upper
198 100
182 18.1 34.5 51.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.357 ng
RT: 16.300 min Scan# 1321
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25





Tgt Ion:248 Resp: 24794
Ion Ratio Lower Upper
248 100
250 100.9 86.2 129.4
141 43.1 36.2 54.2



#22
Hexachlorobenzene
Concen: 0.379 ng
RT: 16.410 min Scan# 1330
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Tgt Ion:284 Resp: 27780
Ion Ratio Lower Upper
284 100
142 33.6 22.2 33.2#
249 30.9 26.6 39.8



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad
6/24/2021 9:53:25 AM



#23
Atrazine
Concen: 0.267 ng
RT: 16.568 min Scan# 1343
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:200 Resp: 16582
Ion Ratio Lower Upper
200 100
173 23.9 21.7 32.5
215 49.2 43.1 64.7

Manual Integrations
APPROVED



mohammad
6/24/2021 9:53:25 AM




#24
Pentachlorophenol
Concen: 0.279 ng
RT: 16.750 min Scan# 1358
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25



Tgt Ion:266 Resp: 7528
Ion Ratio Lower Upper
266 100
264 63.0 51.0 76.4
268 64.0 50.6 76.0









#25
Phenanthrene
Concen: 0.383 ng
RT: 17.140 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:178 Resp: 125592
Ion Ratio Lower Upper
178 100
176 18.7 15.6 23.4
179 15.6 12.1 18.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM



#26
Anthracene
Concen: 0.343 ng
RT: 17.225 min Scan# 1397
Delta R.T. -0.013 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25


Tgt Ion:178 Resp: 110041
Ion Ratio Lower Upper
178 100
176 18.2 15.1 22.7
179 15.3 12.2 18.2





#27
Fluoranthene-d10
Concen: 0.338 ng
RT: 19.138 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:212 Resp: 114483
Ion Ratio Lower Upper
212 100
106 14.2 7.0 10.6#
104 7.8 4.1 6.1#

Manual Integrations
APPROVED



mohammad
6/24/2021 9:53:25 AM




#28
Fluoranthene
Concen: 0.377 ng
RT: 19.163 min Scan# 1693
Delta R.T. -0.005 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Tgt Ion:202 Resp: 139904
Ion Ratio Lower Upper
202 100
101 11.8 6.6 10.0#
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.292 min Scan# 1984
Delta R.T. -0.011 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25





Tgt Ion:240 Resp: 96656
Ion Ratio Lower Upper
240 100
120 11.8 4.2 6.4#
236 26.6 21.4 32.2



#30
Pyrene
Concen: 0.370 ng
RT: 19.525 min Scan# 1766
Delta R.T. -0.010 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Tgt Ion:202 Resp: 141494
Ion Ratio Lower Upper
202 100
200 20.5 16.9 25.3
203 17.4 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad
6/24/2021 9:53:25 AM



#31
Terphenyl-d14
Concen: 0.364 ng
RT: 19.739 min Scan# 1809
Delta R.T. -0.005 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:244 Resp: 87283
Ion Ratio Lower Upper
244 100
212 7.7 6.3 9.5
122 13.5 6.2 9.2#

Manual Integrations
APPROVED



mohammad
6/24/2021 9:53:25 AM




#32
Benzo(a)anthracene
Concen: 0.347 ng
RT: 21.281 min Scan# 1983
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Tgt Ion:228 Resp: 126601
Ion Ratio Lower Upper
228 100
226 26.0 23.0 34.6
229 19.8 15.7 23.5





#33
Chrysene
Concen: 0.384 ng
RT: 21.334 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:228 Resp: 133899
Ion Ratio Lower Upper
228 100
226 28.7 25.0 37.6
229 20.0 15.9 23.9

Manual Integrations
APPROVED



mohammad
6/24/2021 9:53:25 AM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.216 ng
RT: 21.228 min Scan# 1978
Delta R.T. -0.000 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Tgt Ion:149 Resp: 54889
Ion Ratio Lower Upper
149 100
167 28.1 28.9 43.3#
279 3.6 4.9 7.3#





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.540 min Scan# 2420
Delta R.T. -0.014 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:264 Resp: 87145
Ion Ratio Lower Upper
264 100
260 24.6 20.6 31.0
265 23.3 21.4 32.2




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.352 ng
RT: 25.799 min Scan# 3080
Delta R.T. -0.014 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:276 Resp: 102444
Ion Ratio Lower Upper
276 100
138 30.4 10.0 15.0#
277 25.0 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.391 ng
RT: 22.869 min Scan# 2224
Delta R.T. -0.007 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:252 Resp: 127937
Ion Ratio Lower Upper
252 100
253 22.3 18.5 27.7
125 12.0 5.8 8.6#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM



#38
Benzo(k)fluoranthene
Concen: 0.420 ng
RT: 22.910 min Scan# 2236
Delta R.T. -0.011 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:252 Resp: 132525
Ion Ratio Lower Upper
252 100
253 22.3 19.1 28.7
125 12.1 5.4 8.0#









#39
Benzo(a)pyrene
Concen: 0.383 ng
RT: 23.444 min Scan# 2392
Delta R.T. -0.011 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:252 Resp: 110132
Ion Ratio Lower Upper
252 100
253 22.8 19.4 29.2
125 17.7 7.4 11.2#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
6/24/2021 9:53:25 AM




#40
Dibenzo(a,h)anthracene
Concen: 0.339 ng
RT: 25.816 min Scan# 3085
Delta R.T. -0.014 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25

Tgt Ion:278 Resp: 79781
Ion Ratio Lower Upper
278 100
139 19.0 9.0 13.4#
279 23.8 21.0 31.4







#41
Benzo(g,h,i)perylene
Concen: 0.333 ng
RT: 26.487 min Scan# 3281
Delta R.T. -0.017 min
Lab File: BN015028.D
Acq: 24 Jun 2021 04:25



Tgt Ion: 276 Resp: 84188
Ion Ratio Lower Upper
276 100
277 23.7 19.7 29.5
138 23.6 9.8 14.6#



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
6/24/2021 9:53:25 AM