

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062422\
 Data File : BN020388.D
 Acq On : 24 Jun 2022 09:23
 Operator : CG/JU
 Sample : N3355-19
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Jun 25 00:54:27 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:17:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	2839	0.400	ng/ul	0.00
4) Naphthalene-d8	10.529	136	9884	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.378	164	6569	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.121	188	14730	0.400	ng/ul	0.00
17) Chrysene-d12	21.313	240	13234	0.400	ng/ul	0.00
23) Perylene-d12	23.593	264	13015	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.201	96	12787	3.888	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.130	152	5306	0.327	ng/ul	0.00
18) Fluoranthene-d10	19.152	212	15337	0.351	ng/ul	0.00
Target Compounds						
					Qvalue	
12) Fluorene	15.372	166	1846	0.062	ng/ul#	14
20) Pyrene	19.514	202	1832	0.029	ng/ul#	1
26) Benzo(a)pyrene	23.444	252	3967	0.072	ng/ul#	72


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062422\
Data File : BN020388.D
Acq On : 24 Jun 2022 09:23
Operator : CG/JU
Sample : N3355-19
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :


Quant Time: Jun 25 00:54:27 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:17:34 2022
Response via : Initial Calibration








#12
Fluorene
Concen: 0.062 ng/ul
RT: 15.372 min Scan# 2687
Delta R.T. -0.058 min
Lab File: BN020388.D
Acq: 24 Jun 2022 09:23

Instrument :
BNA_N
ClientSampleId :







Tgt Ion:166 Resp: 1846
Ion Ratio Lower Upper
166 100
165 8.4 78.2 117.2#
167 36.1 11.3 16.9#



#20
Pyrene
Concen: 0.029 ng/ul
RT: 19.514 min Scan# 3633
Delta R.T. -0.030 min
Lab File: BN020388.D
Acq: 24 Jun 2022 09:23

Tgt Ion:202 Resp: 1832
Ion Ratio Lower Upper
202 100
101 165.0 11.6 17.4#
100 1183.1 9.2 13.8#





#26

Benzo(a)pyrene

Concen: 0.072 ng/ul

RT: 23.444 min Scan# 41

Delta R.T. -0.049 min


Lab File: BN020388.D

Acq: 24 Jun 2022 09:23

Instrument :

BNA_N

ClientSampleId :



Tgt Ion:252 Resp: 3967

Ion Ratio Lower Upper

252 100

253 41.5 21.0 31.4#

125 36.0 18.8 28.2#

