

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062422\
 Data File : BN020404.D
 Acq On : 24 Jun 2022 19:41
 Operator : CG/JU
 Sample : PB145731BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK731

Quant Time: Jun 25 00:55:29 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:17:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	2773	0.400	ng/u1	0.00
4) Naphthalene-d8	10.535	136	9615	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.382	164	6505	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.120	188	14377	0.400	ng/u1	0.00
17) Chrysene-d12	21.309	240	11847	0.400	ng/u1	0.00
23) Perylene-d12	23.589	264	11543	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.201	96	21563	6.712	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.130	152	5531	0.351	ng/u1	0.00
18) Fluoranthene-d10	19.151	212	13740	0.352	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062422\
Data File : BN020404.D
Acq On : 24 Jun 2022 19:41
Operator : CG/JU
Sample : PB145731BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK731

Quant Time: Jun 25 00:55:29 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:17:34 2022
Response via : Initial Calibration

