

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062422\
 Data File : BN020413.D
 Acq On : 25 Jun 2022 01:09
 Operator : CG/JU
 Sample : N3358-20
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EW4T3

Quant Time: Jun 25 03:53:20 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 00:17:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.750	152	3087	0.400	ng/u1	0.00
4) Naphthalene-d8	10.534	136	10775	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.380	164	7369	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.119	188	16969	0.400	ng/u1	0.00
17) Chrysene-d12	21.306	240	15421	0.400	ng/u1	0.00
23) Perylene-d12	23.586	264	15221	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.201	96	17470	4.885	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.128	152	5564	0.315	ng/u1	0.00
18) Fluoranthene-d10	19.149	212	17679	0.348	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062422\
Data File : BN020413.D
Acq On : 25 Jun 2022 01:09
Operator : CG/JU
Sample : N3358-20
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EW4T3

Quant Time: Jun 25 03:53:20 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN062422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 00:17:34 2022
Response via : Initial Calibration

