

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062424\
 Data File : BN032232.D
 Acq On : 25 Jun 2024 08:06
 Operator : MA/JU
 Sample : PB161513BL
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK513

Quant Time: Jun 25 08:39:03 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 20 10:49:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.628	152	277397	20.000	ng/u1	0.00
20) Naphthalene-d8	10.404	136	1061757	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.269	164	616042	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.022	188	1257982	20.000	ng/u1	-0.01
79) Chrysene-d12	21.257	240	1211265	20.000	ng/u1	-0.01
88) Perylene-d12	24.168	264	1400646	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.164	96	37553	5.601	ng/uL	0.00
4) Pyridine-d5	3.564	84	488459	25.879	ng/u1	0.00
7) Phenol-d5	6.811	99	638325	26.842	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.969	67	345420	24.833	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.163	132	557316	30.714	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	504572	26.029	ng/u1	0.00
21) Nitrobenzene-d5	8.787	128	256640	31.613	ng/u1	0.00
24) 2-Nitrophenol-d4	9.499	143	279346	33.311	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.034	165	463554	29.609	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.551	131	693358	29.829	ng/u1	-0.01
46) Dimethylphthalate-d6	13.687	166	1361204	31.717	ng/u1	-0.01
49) Acenaphthylene-d8	13.963	160	1595196	32.325	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.504	143	237520	27.641	ng/u1	0.00
60) Fluorene-d10	15.269	176	1181006	32.523	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.410	200	167028	26.415	ng/u1	0.00
73) Anthracene-d10	17.122	188	1874205	34.800	ng/u1	0.00
81) Pyrene-d10	19.422	212	2126500	32.807	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.974	264	2245077	33.568	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062424\
Data File : BN032232.D
Acq On : 25 Jun 2024 08:06
Operator : MA/JU
Sample : PB161513BL
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK513

Quant Time: Jun 25 08:39:03 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN052924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 20 10:49:40 2024
Response via : Initial Calibration

