

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062425\
 Data File : BN037368.D
 Acq On : 24 Jun 2025 10:40
 Operator : RC/JU
 Sample : PB168430BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 24 12:54:41 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN062025.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jun 19 14:59:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.557	152	97959	20.000	ng/ul	0.00
20) Naphthalene-d8	10.340	136	311114	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.216	164	211315	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.963	188	418334	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.198	240	389617	20.000	ng/ul	0.00
88) Perylene-d12	24.039	264	415473	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.058	96	15653	8.536	ng/uL	0.00
4) Pyridine-d5	3.469	84	192892	39.372	ng/ul	0.00
7) Phenol-d5	6.740	99	223518	39.573	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.904	67	139829	39.728	ng/ul	0.00
11) 2-Chlorophenol-d4	7.093	132	192532	38.452	ng/ul	0.00
15) 4-Methylphenol-d8	8.269	113	176356	38.140	ng/ul	0.00
21) Nitrobenzene-d5	8.716	128	87962	38.582	ng/ul	0.00
24) 2-Nitrophenol-d4	9.434	143	104318	37.515	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.963	165	213882	35.460	ng/ul	0.00
31) 4-Chloroaniline-d4	10.481	131	201133	33.180	ng/ul	0.00
46) Dimethylphthalate-d6	13.633	166	591866	37.970	ng/ul	0.00
49) Acenaphthylene-d8	13.904	160	667102	38.902	ng/ul	0.00
54) 4-Nitrophenol-d4	14.422	143	62612	27.047	ng/ul	0.00
60) Fluorene-d10	15.216	176	514280	36.507	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.339	200	87703	29.363	ng/ul	0.00
73) Anthracene-d10	17.063	188	773815	37.432	ng/ul	0.00
81) Pyrene-d10	19.368	212	885660	41.792	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.845	264	824871	38.206	ng/ul	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062425\
Data File : BN037368.D
Acq On : 24 Jun 2025 10:40
Operator : RC/JU
Sample : PB168430BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Jun 24 12:54:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN062025.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jun 19 14:59:41 2025
Response via : Initial Calibration

