

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062620\
 Data File : BN011056.D
 Acq On : 26 Jun 2020 11:21
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 26 12:01:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 26 11:59:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.538	0.553	-2.8	87	0.00
3	n-Nitrosodimethylamine	0.300	0.285	5.0	81	0.00
4 S	2-Fluorophenol	1.127	1.165	-3.4	88	0.00
5 S	Phenol-d6	1.335	1.375	-3.0	87	0.00
6	bis(2-Chloroethyl)ether	1.153	1.232	-6.9	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.364	0.332	8.8	77	0.00
9	Naphthalene	1.215	1.180	2.9	94	0.00
10	Hexachlorobutadiene	0.293	0.284	3.1	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.706	0.675	4.4	86	0.00
12	2-Methylnaphthalene	0.814	0.775	4.8	86	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.176	0.168	4.5	91	0.00
15 S	2-Fluorobiphenyl	1.866	1.918	-2.8	94	0.00
16	Acenaphthylene	2.016	2.028	-0.6	97	0.00
17	Acenaphthene	1.354	1.377	-1.7	95	0.00
18	Fluorene	1.727	1.704	1.3	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.067	0.013	80.6#	21#	0.00
21	4-Bromophenyl-phenylether	0.277	0.275	0.7	95	0.00
22	Hexachlorobenzene	0.306	0.310	-1.3	95	0.00
23	Atrazine	0.204	0.191	6.4	94	0.00
24	Pentachlorophenol	0.097	0.086	11.3	87	0.00
25	Phenanthrene	1.348	1.316	2.4	93	0.00
26	Anthracene	1.119	1.093	2.3	96	0.00
27 SURR	Fluoranthene-d10	1.265	1.220	3.6	95	0.00
28	Fluoranthene	1.541	1.450	5.9	90	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
30	Pyrene	1.539	1.549	-0.6	96	0.00
31 S	Terphenyl-d14	1.074	1.102	-2.6	96	0.00
32	Benzo(a)anthracene	1.386	1.423	-2.7	96	0.00
33	Chrysene	1.511	1.596	-5.6	101	0.00
34	Bis(2-ethylhexyl)phthalate	0.557	0.599	-7.5	102	0.00
35	Indeno(1,2,3-cd)pyrene	1.680	1.814	-8.0	98	0.00
36 I	Perylene-d12	1.000	1.000	0.0	87	0.00
37	Benzo(b)fluoranthene	1.582	1.581	0.1	98	0.00
38	Benzo(k)fluoranthene	1.645	1.653	-0.5	102	0.00
39 C	Benzo(a)pyrene	1.405	1.394	0.8	85	0.00
40	Dibenzo(a,h)anthracene	1.515	1.490	1.7	97	0.00
41	Benzo(g,h,i)perylene	1.569	1.542	1.7	100	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062620\
Data File : BN011056.D
Acq On : 26 Jun 2020 11:21
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 26 12:01:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 26 11:59:47 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0