

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062620\
 Data File : BN011063.D
 Acq On : 26 Jun 2020 17:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 26 17:51:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 26 11:59:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.538	0.578	-7.4	86	0.00
3	n-Nitrosodimethylamine	0.300	0.289	3.7	78	0.00
4 S	2-Fluorophenol	1.127	1.111	1.4	79	0.00
5 S	Phenol-d6	1.335	1.311	1.8	78	0.00
6	bis(2-Chloroethyl)ether	1.153	1.177	-2.1	81	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	84	0.00
8 S	Nitrobenzene-d5	0.364	0.351	3.6	71	0.00
9	Naphthalene	1.215	1.217	-0.2	86	0.00
10	Hexachlorobutadiene	0.293	0.300	-2.4	83	0.00
11 SURR	2-Methylnaphthalene-d10	0.706	0.693	1.8	78	0.00
12	2-Methylnaphthalene	0.814	0.810	0.5	79	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
14 S	2,4,6-Tribromophenol	0.176	0.176	0.0	85	0.00
15 S	2-Fluorobiphenyl	1.866	1.958	-4.9	85	0.00
16	Acenaphthylene	2.016	2.064	-2.4	88	0.00
17	Acenaphthene	1.354	1.384	-2.2	85	0.00
18	Fluorene	1.727	1.740	-0.8	84	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
20	4,6-Dinitro-2-methylphenol	0.067	0.028	58.2#	39#	0.00
21	4-Bromophenyl-phenylether	0.277	0.271	2.2	82	0.00
22	Hexachlorobenzene	0.306	0.322	-5.2	87	0.00
23	Atrazine	0.204	0.197	3.4	85	0.00
24	Pentachlorophenol	0.097	0.083	14.4	74	-0.01
25	Phenanthrene	1.348	1.347	0.1	84	0.00
26	Anthracene	1.119	1.103	1.4	85	0.00
27 SURR	Fluoranthene-d10	1.265	1.254	0.9	86	0.00
28	Fluoranthene	1.541	1.488	3.4	81	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
30	Pyrene	1.539	1.556	-1.1	86	0.00
31 S	Terphenyl-d14	1.074	1.103	-2.7	85	0.00
32	Benzo(a)anthracene	1.386	1.400	-1.0	84	-0.01
33	Chrysene	1.511	1.538	-1.8	87	0.00
34	Bis(2-ethylhexyl)phthalate	0.557	0.591	-6.1	89	-0.01
35	Indeno(1,2,3-cd)pyrene	1.680	1.811	-7.8	87	0.00
36 I	Perylene-d12	1.000	1.000	0.0	76	0.00
37	Benzo(b)fluoranthene	1.582	1.555	1.7	84	0.00
38	Benzo(k)fluoranthene	1.645	1.744	-6.0	94	0.00
39 C	Benzo(a)pyrene	1.405	1.426	-1.5	76	0.00
40	Dibenzo(a,h)anthracene	1.515	1.518	-0.2	86	0.00
41	Benzo(g,h,i)perylene	1.569	1.566	0.2	89	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN062620\
Data File : BN011063.D
Acq On : 26 Jun 2020 17:18
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 26 17:51:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN062320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 26 11:59:47 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0