

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062724\
 Data File : BN032281.D
 Acq On : 27 Jun 2024 15:29
 Operator : MA/JU
 Sample : P2909-16
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4DN6

Quant Time: Jun 27 15:57:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 27 13:23:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.620	152	7848	0.400	ng/ul	0.00
4) Naphthalene-d8	10.397	136	26484	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.261	164	17264	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.015	188	38841	0.400	ng/ul	0.00
17) Chrysene-d12	21.222	240	28313	0.400	ng/ul	0.00
23) Perylene-d12	23.443	264	25653	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.152	96	4300	0.493	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.992	152	12360	0.311	ng/ul	0.00
18) Fluoranthene-d10	19.053	212	33647	0.395	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.441	128	331279	4.727	ng/ul	99
10) Acenaphthylene	13.983	152	573878	7.993	ng/ul	99
11) Acenaphthene	14.326	153	837318	15.026	ng/ul	99
12) Fluorene	15.320	166	180730	2.722	ng/ul	98
15) Phenanthrene	17.057	178	1601575	15.040	ng/ul	100
16) Anthracene	17.150	178	1359127	14.309	ng/ul	100
19) Fluoranthene	19.086	202	943363	8.750	ng/ul	99
20) Pyrene	19.448	202	1124764	9.983	ng/ul	98
21) Benzo(a)anthracene	21.204	228	419586	4.106	ng/ul	100
22) Chrysene	21.257	228	693105	6.180	ng/ul	99
24) Benzo(b)fluoranthene	22.773	252	768592	6.896	ng/ul	92
25) Benzo(k)fluoranthene	22.817	252	527297	4.869	ng/ul	94
26) Benzo(a)pyrene	23.346	252	84293	1.087	ng/ul	96
27) Indeno(1,2,3-cd)pyrene	25.678	276	584400	4.813	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.688	278	472467	5.237	ng/ul	95
29) Benzo(g,h,i)perylene	26.366	276	410067	4.240	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062724\
Data File : BN032281.D
Acq On : 27 Jun 2024 15:29
Operator : MA/JU
Sample : P2909-16
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A4DN6

Quant Time: Jun 27 15:57:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN062724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 13:23:38 2024
Response via : Initial Calibration

