

Data Path : \\74.0.250.170\TERASTORAGE\SVOASRV\HPCHEM1\BNA_N\DATA\BN062819\
 Data File : BN006626.D
 Acq On : 28 Jun 2019 22:45
 Operator : HP/JU
 Sample : K3427-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 KY038PS037-190617

Quant Time: Jun 29 01:39:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 13:47:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	3256	0.40	ng	0.00
7) Naphthalene-d8	10.40	136	13408	0.40	ng	0.00
13) Acenaphthene-d10	14.28	164	8949	0.40	ng	0.00
19) Phenanthrene-d10	17.04	188	20003	0.40	ng	0.00
27) Chrysene-d12	21.26	240	21575	0.40	ng	-0.01
34) Perylene-d12	23.48	264	20693	0.40	ng	-0.04
System Monitoring Compounds						
4) 2-Fluorophenol	5.24	112	1448	0.18	ng	0.00
5) Phenol-d6	6.84	99	1102	0.10	ng	0.00
8) Nitrobenzene-d5	8.81	82	3843	0.43	ng	0.01
11) 2-Methylnaphthalene-d10	12.01	152	5651	0.27	ng	0.00
14) 2,4,6-Tribromophenol	15.79	330	1357	0.32	ng	0.00
15) 2-Fluorobiphenyl	12.88	172	19591	0.59	ng	-0.01
25) Fluoranthene-d10	19.08	212	17184	0.29	ng	0.00
29) Terphenyl-d14	19.68	244	25228	0.56	ng	0.00
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.15	149	2761	0.078	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\TERASTORAGE\SVOASRV\HPCHEM1\BNA_N\DATA\BN062819\
Data File : BN006626.D
Acq On : 28 Jun 2019 22:45
Operator : HP/JU
Sample : K3427-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
KY038PS037-190617

Quant Time: Jun 29 01:39:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 13:47:56 2019
Response via : Initial Calibration

