

Data Path : \\74.0.250.170\TERASTORAGE\SVOASRV\HPCHEM1\BNA_N\DATA\BN062819\
 Data File : BN006635.D
 Acq On : 29 Jun 2019 03:56
 Operator : HP/JU
 Sample : K3486-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 W-2

Quant Time: Jun 29 04:37:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 29 01:19:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	4905	0.40	ng	0.00
7) Naphthalene-d8	10.40	136	22046	0.40	ng	0.00
13) Acenaphthene-d10	14.27	164	14826	0.40	ng	-0.01
19) Phenanthrene-d10	17.03	188	32864	0.40	ng	-0.01
27) Chrysene-d12	21.25	240	29736	0.40	ng	-0.02
34) Perylene-d12	23.47	264	27126	0.40	ng	-0.05

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.25	112	1732	0.14	ng	0.00
5) Phenol-d6	6.81	99	1683	0.11	ng	-0.02
8) Nitrobenzene-d5	8.79	82	5782	0.39	ng	0.00
11) 2-Methylnaphthalene-d10	12.00	152	13177	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.78	330	3051	0.44	ng	-0.01
15) 2-Fluorobiphenyl	12.88	172	21114	0.38	ng	-0.01
25) Fluoranthene-d10	19.07	212	35001	0.36	ng	0.00
29) Terphenyl-d14	19.67	244	31266	0.50	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.17	88	210	0.031	ng	# 1
22) Pentachlorophenol	16.71	266	260	0.241	ng	97
24) Anthracene	17.16	178	2684	0.033	ng	# 83
26) Fluoranthene	19.10	202	2275	0.020	ng	# 42
32) Bis(2-ethylhexyl)phthalate	21.15	149	9983	0.204	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\TERASTORAGE\SVOASRV\HPCHEM1\BNA_N\DATA\BN062819\
Data File : BN006635.D
Acq On : 29 Jun 2019 03:56
Operator : HP/JU
Sample : K3486-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
W-2

Quant Time: Jun 29 04:37:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 29 01:19:02 2019
Response via : Initial Calibration

