

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
 Data File : BN006611.D
 Acq On : 28 Jun 2019 13:47
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062819

Quant Time: Jun 28 18:38:04 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 18:31:55 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.554	0.598	-7.9	113	0.00
3	n-Nitrosodimethylamine	0.435	0.402	7.6	102	0.00
4 S	2-Fluorophenol	0.981	0.950	3.2	99	0.00
5 S	Phenol-d6	1.299	1.226	5.6	98	0.00
6	bis(2-Chloroethyl)ether	1.158	1.183	-2.2	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.269	0.247	8.2	98	0.01
9	Naphthalene	1.051	1.079	-2.7	100	0.00
10	Hexachlorobutadiene	0.216	0.224	-3.7	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.630	0.642	-1.9	99	0.00
12	2-Methylnaphthalene	0.726	0.749	-3.2	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.188	0.171	9.0	100	0.00
15 S	2-Fluorobiphenyl	1.482	1.499	-1.1	100	0.00
16	Acenaphthylene	1.907	1.850	3.0	98	0.01
17	Acenaphthene	1.256	1.311	-4.4	101	0.00
18	Fluorene	1.585	1.629	-2.8	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.222	0.223	-0.5	101	0.00
21	Hexachlorobenzene	0.280	0.291	-3.9	100	0.00
22	Pentachlorophenol	0.045	0.037	17.8	103	0.00
23	Phenanthrene	1.137	1.146	-0.8	100	0.01
24	Anthracene	0.987	0.938	5.0	98	0.01
25 SURR	Fluoranthene-d10	1.175	1.169	0.5	98	0.00
26	Fluoranthene	1.382	1.403	-1.5	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	0.01
28	Pyrene	1.400	1.454	-3.9	98	0.00
29 S	Terphenyl-d14	0.838	0.862	-2.9	99	0.00
30	Benzo(a)anthracene	1.308	1.264	3.4	95	0.00
31	Chrysene	1.410	1.467	-4.0	95	0.01
32	Bis(2-ethylhexyl)phthalate	0.658	0.576	12.5	94	0.01
33	Indeno(1,2,3-cd)pyrene	1.427	1.409	1.3	82	0.02
34 I	Perylene-d12	1.000	1.000	0.0	86	0.01
35	Benzo(b)fluoranthene	1.361	1.435	-5.4	90	0.01
36	Benzo(k)fluoranthene	1.450	1.558	-7.4	90	0.01
37 C	Benzo(a)pyrene	1.272	1.340	-5.3	87	0.01
38	Dibenzo(a,h)anthracene	1.111	1.124	-1.2	82	0.02
39	Benzo(g,h,i)perylene	1.109	1.127	-1.6	82	0.02

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
Data File : BN006611.D
Acq On : 28 Jun 2019 13:47
Operator : HP/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN062819

Quant Time: Jun 28 18:38:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 18:31:55 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			