

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
 Data File : BN006611.D
 Acq On : 28 Jun 2019 13:47
 Operator : HP/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN062819

Quant Time: Jun 28 18:38:04 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 18:31:55 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.432	-8.0	113	0.00
3	n-Nitrosodimethylamine	0.400	0.369	7.8	102	0.00
4 S	2-Fluorophenol	0.400	0.387	3.3	99	0.00
5 S	Phenol-d6	0.400	0.378	5.5	98	0.00
6	bis(2-Chloroethyl)ether	0.400	0.409	-2.2	96	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	99	0.00
8 S	Nitrobenzene-d5	0.400	0.368	8.0	98	0.01
9	Naphthalene	0.400	0.411	-2.7	100	0.00
10	Hexachlorobutadiene	0.400	0.415	-3.7	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.408	-2.0	99	0.00
12	2-Methylnaphthalene	0.400	0.412	-3.0	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.400	0.362	9.5	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.405	-1.3	100	0.00
16	Acenaphthylene	0.400	0.388	3.0	98	0.01
17	Acenaphthene	0.400	0.417	-4.2	101	0.00
18	Fluorene	0.400	0.411	-2.7	100	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.400	0.403	-0.8	101	0.00
21	Hexachlorobenzene	0.400	0.415	-3.7	100	0.00
22	Pentachlorophenol	0.400	0.422	-5.5	103	0.00
23	Phenanthrene	0.400	0.403	-0.8	100	0.01
24	Anthracene	0.400	0.380	5.0	98	0.01
25 SURR	Fluoranthene-d10	0.400	0.398	0.5	98	0.00
26	Fluoranthene	0.400	0.406	-1.5	98	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	96	0.01
28	Pyrene	0.400	0.416	-4.0	98	0.00
29 S	Terphenyl-d14	0.400	0.411	-2.7	99	0.00
30	Benzo(a)anthracene	0.400	0.387	3.3	95	0.00
31	Chrysene	0.400	0.416	-4.0	95	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.350	12.5	94	0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.395	1.3	82	0.02
34 I	Perylene-d12	0.400	0.400	0.0	86	0.01
35	Benzo(b)fluoranthene	0.400	0.422	-5.5	90	0.01
36	Benzo(k)fluoranthene	0.400	0.430	-7.5	90	0.01
37 C	Benzo(a)pyrene	0.400	0.421	-5.2	87	0.01
38	Dibenzo(a,h)anthracene	0.400	0.405	-1.3	82	0.02
39	Benzo(g,h,i)perylene	0.400	0.406	-1.5	82	0.02

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
Data File : BN006611.D
Acq On : 28 Jun 2019 13:47
Operator : HP/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN062819

Quant Time: Jun 28 18:38:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 18:31:55 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				