

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
 Data File : BN006620.D
 Acq On : 28 Jun 2019 19:17
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 01 04:56:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 13:47:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.554	0.538	2.9	102	0.00
3	n-Nitrosodimethylamine	0.435	0.400	8.0	102	0.00
4 S	2-Fluorophenol	0.981	0.844	14.0	88	0.00
5 S	Phenol-d6	1.299	1.152	11.3	92	0.00
6	bis(2-Chloroethyl)ether	1.158	1.091	5.8	89	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.269	0.236	12.3	92	0.01
9	Naphthalene	1.051	1.085	-3.2	99	0.00
10	Hexachlorobutadiene	0.216	0.229	-6.0	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.630	0.639	-1.4	97	0.00
12	2-Methylnaphthalene	0.726	0.746	-2.8	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.188	0.162	13.8	94	0.00
15 S	2-Fluorobiphenyl	1.482	1.488	-0.4	99	0.00
16	Acenaphthylene	1.907	1.835	3.8	96	0.00
17	Acenaphthene	1.256	1.311	-4.4	100	0.00
18	Fluorene	1.585	1.618	-2.1	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.222	0.219	1.4	98	0.00
21	Hexachlorobenzene	0.280	0.293	-4.6	100	0.00
22	Pentachlorophenol	0.045	0.034	24.4	93	0.00
23	Phenanthrene	1.137	1.149	-1.1	99	0.00
24	Anthracene	0.987	0.932	5.6	96	0.01
25 SURR	Fluoranthene-d10	1.175	1.166	0.8	97	0.00
26	Fluoranthene	1.382	1.405	-1.7	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	-0.02
28	Pyrene	1.400	1.450	-3.6	96	0.00
29 S	Terphenyl-d14	0.838	0.853	-1.8	97	0.00
30	Benzo(a)anthracene	1.308	1.306	0.2	97	-0.02
31	Chrysene	1.410	1.413	-0.2	91	-0.02
32	Bis(2-ethylhexyl)phthalate	0.658	0.540	17.9	87	-0.02
33	Indeno(1,2,3-cd)pyrene	1.427	1.464	-2.6	84	-0.04
34 I	Perylene-d12	1.000	1.000	0.0	88	-0.04
35	Benzo(b)fluoranthene	1.361	1.472	-8.2	95	-0.04
36	Benzo(k)fluoranthene	1.450	1.530	-5.5	91	-0.04
37 C	Benzo(a)pyrene	1.272	1.343	-5.6	89	-0.04
38	Dibenzo(a,h)anthracene	1.111	1.127	-1.4	84	-0.04
39	Benzo(g,h,i)perylene	1.109	1.115	-0.5	83	-0.04

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
Data File : BN006620.D
Acq On : 28 Jun 2019 19:17
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 01 04:56:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 13:47:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		