

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
 Data File : BN006620.D
 Acq On : 28 Jun 2019 19:17
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 01 04:56:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 13:47:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.389	2.8	102	0.00
3	n-Nitrosodimethylamine	0.400	0.368	8.0	102	0.00
4 S	2-Fluorophenol	0.400	0.344	14.0	88	0.00
5 S	Phenol-d6	0.400	0.355	11.3	92	0.00
6	bis(2-Chloroethyl)ether	0.400	0.377	5.8	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	98	0.00
8 S	Nitrobenzene-d5	0.400	0.351	12.3	92	0.01
9	Naphthalene	0.400	0.413	-3.2	99	0.00
10	Hexachlorobutadiene	0.400	0.425	-6.2	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.406	-1.5	97	0.00
12	2-Methylnaphthalene	0.400	0.411	-2.7	97	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.400	0.344	14.0	94	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	99	0.00
16	Acenaphthylene	0.400	0.385	3.8	96	0.00
17	Acenaphthene	0.400	0.417	-4.2	100	0.00
18	Fluorene	0.400	0.408	-2.0	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.400	0.394	1.5	98	0.00
21	Hexachlorobenzene	0.400	0.419	-4.7	100	0.00
22	Pentachlorophenol	0.400	0.403	-0.8	93	0.00
23	Phenanthrene	0.400	0.404	-1.0	99	0.00
24	Anthracene	0.400	0.378	5.5	96	0.01
25 SURR	Fluoranthene-d10	0.400	0.397	0.8	97	0.00
26	Fluoranthene	0.400	0.407	-1.7	97	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	95	-0.02
28	Pyrene	0.400	0.414	-3.5	96	0.00
29 S	Terphenyl-d14	0.400	0.407	-1.7	97	0.00
30	Benzo(a)anthracene	0.400	0.399	0.3	97	-0.02
31	Chrysene	0.400	0.401	-0.3	91	-0.02
32	Bis(2-ethylhexyl)phthalate	0.400	0.328	18.0	87	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.410	-2.5	84	-0.04
34 I	Perylene-d12	0.400	0.400	0.0	88	-0.04
35	Benzo(b)fluoranthene	0.400	0.433	-8.2	95	-0.04
36	Benzo(k)fluoranthene	0.400	0.422	-5.5	91	-0.04
37 C	Benzo(a)pyrene	0.400	0.422	-5.5	89	-0.04
38	Dibenzo(a,h)anthracene	0.400	0.406	-1.5	84	-0.04
39	Benzo(g,h,i)perylene	0.400	0.402	-0.5	83	-0.04

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
Data File : BN006620.D
Acq On : 28 Jun 2019 19:17
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 01 04:56:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 13:47:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				