

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
 Data File : BN006639.D
 Acq On : 29 Jun 2019 06:48
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 01 08:14:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 13:47:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	0.00
2	1,4-Dioxane	0.554	0.528	4.7	125	0.00
3	n-Nitrosodimethylamine	0.435	0.374	14.0	119	0.00
4 S	2-Fluorophenol	0.981	0.926	5.6	121	0.00
5 S	Phenol-d6	1.299	1.285	1.1	128	0.00
6	bis(2-Chloroethyl)ether	1.158	1.188	-2.6	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
8 S	Nitrobenzene-d5	0.269	0.247	8.2	122	0.00
9	Naphthalene	1.051	1.079	-2.7	125	0.00
10	Hexachlorobutadiene	0.216	0.223	-3.2	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.630	0.651	-3.3	125	0.00
12	2-Methylnaphthalene	0.726	0.756	-4.1	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	126	-0.01
14 S	2,4,6-Tribromophenol	0.188	0.176	6.4	131	0.00
15 S	2-Fluorobiphenyl	1.482	1.479	0.2	126	-0.01
16	Acenaphthylene	1.907	1.877	1.6	126	0.00
17	Acenaphthene	1.256	1.295	-3.1	127	0.00
18	Fluorene	1.585	1.603	-1.1	126	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	125	-0.01
20	4-Bromophenyl-phenylether	0.222	0.219	1.4	125	-0.01
21	Hexachlorobenzene	0.280	0.288	-2.9	125	0.00
22	Pentachlorophenol	0.045	0.039	13.3	138	-0.01
23	Phenanthrene	1.137	1.146	-0.8	125	0.00
24	Anthracene	0.987	0.949	3.9	124	0.00
25 SURR	Fluoranthene-d10	1.175	1.158	1.4	123	0.00
26	Fluoranthene	1.382	1.391	-0.7	123	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	115	-0.01
28	Pyrene	1.400	1.519	-8.5	122	0.00
29 S	Terphenyl-d14	0.838	0.899	-7.3	123	0.00
30	Benzo(a)anthracene	1.308	1.339	-2.4	120	-0.02
31	Chrysene	1.410	1.449	-2.8	112	-0.02
32	Bis(2-ethylhexyl)phthalate	0.658	0.664	-0.9	129	-0.01
33	Indeno(1,2,3-cd)pyrene	1.427	1.330	6.8	92	-0.04
34 I	Perylene-d12	1.000	1.000	0.0	99	-0.04
35	Benzo(b)fluoranthene	1.361	1.454	-6.8	105	-0.03
36	Benzo(k)fluoranthene	1.450	1.575	-8.6	105	-0.04
37 C	Benzo(a)pyrene	1.272	1.349	-6.1	101	-0.04
38	Dibenzo(a,h)anthracene	1.111	1.098	1.2	92	-0.04
39	Benzo(g,h,i)perylene	1.109	1.083	2.3	91	-0.04

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
Data File : BN006639.D
Acq On : 29 Jun 2019 06:48
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 01 08:14:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 13:47:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		