

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
 Data File : BN006650.D
 Acq On : 29 Jun 2019 13:08
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 01 08:16:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 13:47:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	147	0.00
2	1,4-Dioxane	0.400	0.358	10.5	141	0.00
3	n-Nitrosodimethylamine	0.400	0.399	0.3	166	0.00
4 S	2-Fluorophenol	0.400	0.439	-9.7	169	0.00
5 S	Phenol-d6	0.400	0.467	-16.8	182	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	142	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	149	0.00
8 S	Nitrobenzene-d5	0.400	0.441	-10.2	176	-0.01
9	Naphthalene	0.400	0.424	-6.0	155	-0.01
10	Hexachlorobutadiene	0.400	0.432	-8.0	156	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.411	-2.7	150	0.00
12	2-Methylnaphthalene	0.400	0.418	-4.5	151	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	132	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.416	-4.0	152	-0.01
15 S	2-Fluorobiphenyl	0.400	0.440	-10.0	145	-0.01
16	Acenaphthylene	0.400	0.448	-12.0	151	0.00
17	Acenaphthene	0.400	0.431	-7.7	139	0.00
18	Fluorene	0.400	0.418	-4.5	136	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	119	-0.01
20	4-Bromophenyl-phenylether	0.400	0.449	-12.2	135	-0.01
21	Hexachlorobenzene	0.400	0.434	-8.5	125	0.00
22	Pentachlorophenol	0.400	0.512	-28.0#	171	-0.02
23	Phenanthrene	0.400	0.431	-7.7	127	0.00
24	Anthracene	0.400	0.436	-9.0	134	0.00
25 SURR	Fluoranthene-d10	0.400	0.377	5.8	111	0.00
26	Fluoranthene	0.400	0.380	5.0	110	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	77	-0.01
28	Pyrene	0.400	0.559	-39.8#	105	0.00
29 S	Terphenyl-d14	0.400	0.570	-42.5#	109	0.00
30	Benzo(a)anthracene	0.400	0.457	-14.2	89	-0.01
31	Chrysene	0.400	0.430	-7.5	78	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.764	-91.0#	164	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.447	-11.7	74	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	67	-0.02
35	Benzo(b)fluoranthene	0.400	0.417	-4.2	69	-0.02
36	Benzo(k)fluoranthene	0.400	0.401	-0.3	65	-0.02
37 C	Benzo(a)pyrene	0.400	0.431	-7.7	69	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.469	-17.2	73	-0.02
39	Benzo(g,h,i)perylene	0.400	0.482	-20.5	75	-0.02

Data Path : \\74.0.250.170\terastorage\svoasrv\HPCHEM1\BNA_N\Data\BN062819\
Data File : BN006650.D
Acq On : 29 Jun 2019 13:08
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 01 08:16:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 13:47:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				