

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062821\
 Data File : BN015165.D
 Acq On : 28 Jun 2021 08:41
 Operator : CG/JU
 Sample : DFTPP253
 Misc :
 ALS Vial : 1 Sample Multiplier: 1



Instrument :
 BNA_N
 ClientSampleId :
 DFTPP253

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN061821MA.M

Title : SVOA CALIBRATION

Last Update : Thu Jun 24 14:43:13 2021



AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.8	95592	PASS
68	69	0.00	2	1.7	1763	PASS
69	198	0.00	100	34.6	100803	PASS
70	69	0.00	2	0.5	527	PASS
127	198	10	80	44.7	130477	PASS
197	198	0.00	2	0.6	1847	PASS
198	198	100	100	100.0	291584	PASS
199	198	5	9	6.6	19251	PASS
275	198	10	60	24.1	70341	PASS
365	198	1	100	2.9	8520	PASS
441	198	0.01	100	14.3	41581	PASS
442	198	50	100	96.0	280000	PASS
443	442	15	24	20.0	55992	PASS

m/z	Abundance
40.00	621.0
44.00	4662.0
48.90	629.0
73.10	331.0
84.00	458.0
207.10	2254.0
208.00	424.0
209.00	329.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

m/z	Abundance
37.00	492.0
38.00	1390.0
39.00	8535.0
40.00	725.0
41.00	552.0
44.00	4180.0
45.00	325.0
49.00	1077.0
50.00	28288.0
51.00	116824.0
52.00	5710.0
55.00	636.0
56.00	3286.0
57.00	8332.0
58.00	363.0
61.00	1465.0
62.00	1517.0
63.00	4044.0
63.90	833.0
65.00	2336.0
68.00	2139.0
69.00	120184.0
70.00	554.0
73.00	1229.0
74.00	11637.0
75.00	18520.0
76.10	6744.0
77.00	139456.0
78.00	10008.0
79.00	9133.0
80.00	6942.0
81.00	9452.0
82.00	2095.0
82.90	2088.0
83.90	584.0
85.00	1607.0
86.00	3119.0
87.00	1545.0
87.90	625.0
91.00	2342.0
92.00	2191.0
93.00	15645.0
94.00	1106.0
96.10	1049.0
97.10	344.0
98.00	11153.0
99.00	10322.0
99.90	930.0
101.00	5180.0
101.90	355.0
103.00	2025.0
104.00	3350.0
105.00	3409.0
106.00	1144.0
107.00	43104.0
108.00	6561.0
109.00	1413.0
110.00	80400.0
111.00	11742.0
112.00	1365.0
113.00	548.0
116.10	2443.0
117.00	32040.0
118.00	2714.0
118.90	416.0
120.00	548.0
122.00	2465.0
122.90	4219.0
124.00	1981.0
124.90	1742.0
126.10	923.0
127.00	149504.0
128.00	11372.0
129.00	56112.0
130.00	4617.0
130.90	898.0
132.00	694.0
132.90	462.0
134.00	1433.0
135.00	4335.0
136.00	1992.0
137.00	2775.0
137.90	457.0
139.90	690.0
141.00	7144.0
142.00	2423.0
142.90	1821.0
143.90	546.0
145.00	378.0
146.00	1267.0
147.00	3676.0
148.00	8380.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

149.00	1693.0
149.90	450.0
151.00	1197.0
151.70	686.0
152.90	1884.0
153.90	1946.0
155.00	4255.0
156.00	5854.0
156.90	1174.0
157.90	1318.0
159.00	953.0
160.00	2326.0
161.00	3477.0
162.00	840.0
163.00	412.0
164.10	353.0
165.00	2602.0
166.00	2118.0
167.00	13766.0
168.00	10052.0
169.00	1660.0
170.20	479.0
170.80	593.0
171.90	1312.0
173.00	1836.0
174.00	3503.0
175.00	5917.0
176.00	1383.0
177.00	2775.0
178.00	886.0
179.00	10996.0
180.00	7458.0
181.00	3452.0
181.90	533.0
182.80	401.0
184.00	742.0
185.00	5083.0
186.00	40360.0
187.00	11949.0
188.10	1515.0
189.00	2495.0
189.80	310.0
191.00	1335.0
192.00	3517.0
193.00	3856.0
194.10	869.0
195.00	382.0
196.00	8929.0
197.00	1893.0
198.00	326208.0
198.90	21752.0
200.00	1734.0
201.60	1314.0
202.90	1975.0
204.00	10055.0
205.00	17680.0
206.00	78616.0
207.00	12653.0
208.00	3015.0
209.00	940.0
210.00	1223.0
210.90	2432.0
211.80	905.0
214.90	749.0
216.00	1562.0
217.00	18312.0
218.00	2797.0
219.10	301.0
221.00	11985.0
223.00	4085.0
224.00	40136.0
225.00	10463.0
226.00	1612.0
227.00	15550.0
228.00	2692.0
229.00	3760.0
230.00	586.0
230.90	1151.0
234.00	1154.0
234.90	1307.0
235.90	638.0
237.00	1617.0
239.00	443.0
239.90	583.0
240.90	945.0
242.00	2313.0
243.10	2393.0
244.00	32480.0
245.00	4520.0
246.00	5899.0
246.90	1190.0
249.00	925.0
251.90	351.0
253.00	835.0
255.00	148864.0
256.00	24808.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

257.00	1639.0
258.00	8658.0
259.00	1774.0
261.00	337.0
265.00	3735.0
265.90	656.0
271.00	413.0
273.00	4579.0
274.00	12751.0
275.00	72264.0
276.00	9558.0
277.00	5211.0
278.00	1166.0
281.00	382.0
283.00	704.0
284.00	439.0
285.00	1061.0
293.00	1131.0
294.00	491.0
295.00	427.0
296.00	17816.0
296.90	2748.0
303.00	2338.0
304.00	618.0
310.00	335.0
314.00	708.0
315.00	1751.0
316.00	1250.0
321.00	652.0
323.00	6571.0
324.00	1042.0
327.00	1269.0
328.00	508.0
332.00	450.0
333.00	559.0
334.00	3694.0
335.00	1187.0
340.90	837.0
346.00	1482.0
352.00	1748.0
353.00	1182.0
354.00	2082.0
355.00	440.0
365.00	8279.0
365.90	1152.0
370.90	462.0
372.00	3207.0
373.00	859.0
383.00	1108.0
389.90	381.0
390.90	305.0
402.00	1317.0
403.00	1913.0
404.00	800.0
421.00	2060.0
422.00	2042.0
423.00	12885.0
424.10	3140.0
425.00	486.0
441.00	37376.0
442.00	232128.0
443.00	50496.0
444.10	4404.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

m/z	Abundance
36.90	318.0
38.00	1259.0
39.00	7474.0
40.00	819.0
40.90	367.0
44.00	4343.0
49.00	1272.0
50.00	26576.0
51.00	106264.0
52.00	5636.0
54.90	635.0
56.00	3103.0
57.00	7365.0
58.10	393.0
61.00	1391.0
62.00	1672.0
63.00	4420.0
64.00	491.0
65.00	2344.0
68.00	1959.0
69.00	115048.0
70.00	536.0
73.00	971.0
74.00	11644.0
75.00	18288.0
76.00	6457.0
77.00	139392.0
78.00	9333.0
79.00	8867.0
80.00	6609.0
81.00	9118.0
81.90	2377.0
83.00	2464.0
84.00	704.0
85.00	1391.0
86.00	2635.0
87.00	1072.0
87.90	558.0
91.00	2237.0
92.00	2442.0
93.00	14326.0
94.00	930.0
96.00	955.0
96.90	385.0
98.00	10438.0
99.00	9739.0
100.00	956.0
101.00	5285.0
102.00	338.0
103.00	2042.0
103.90	2961.0
104.90	3275.0
106.00	1021.0
107.00	42272.0
108.00	6329.0
108.90	1138.0
110.00	82096.0
111.00	11728.0
111.90	1498.0
112.90	503.0
114.80	311.0
116.00	2198.0
117.00	32184.0
118.00	2588.0
119.10	344.0
120.00	483.0
122.00	2724.0
123.00	4210.0
124.00	1836.0
125.00	1792.0
127.00	149888.0
128.00	11207.0
129.00	54096.0
130.00	4673.0
130.90	1023.0
131.90	465.0
133.90	1804.0
135.00	4579.0
136.00	2167.0
137.00	2180.0
138.00	429.0
139.00	335.0
139.90	860.0
141.00	6707.0
142.00	2740.0
143.00	1885.0
144.00	502.0
144.90	481.0
145.90	1198.0
147.00	3762.0
148.00	8021.0
149.00	1773.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

150.00	620.0
151.10	1049.0
151.80	632.0
153.00	2068.0
154.00	1914.0
155.00	3989.0
156.00	6829.0
157.00	1240.0
157.90	1379.0
159.00	937.0
160.00	1761.0
161.00	3764.0
161.90	1216.0
162.90	462.0
163.90	334.0
165.00	2558.0
166.00	2398.0
167.00	14465.0
168.00	9463.0
169.00	1518.0
169.90	605.0
170.90	667.0
171.90	1582.0
172.90	2049.0
174.00	3050.0
175.00	6103.0
176.00	1684.0
177.00	2797.0
178.00	942.0
179.00	11240.0
180.00	8057.0
181.00	3306.0
182.00	522.0
183.00	320.0
183.90	825.0
185.00	5429.0
186.00	43664.0
187.00	11702.0
188.00	1477.0
189.00	2675.0
191.00	1500.0
192.00	3756.0
193.00	4150.0
193.90	844.0
195.00	626.0
196.00	9039.0
197.00	2240.0
198.00	337216.0
199.00	21640.0
200.00	1804.0
201.50	1484.0
203.00	2155.0
204.00	10746.0
205.00	19864.0
206.00	84296.0
207.00	12936.0
208.00	3157.0
209.00	1214.0
210.90	3051.0
212.00	490.0
215.00	957.0
216.00	1536.0
217.00	21416.0
217.90	2632.0
219.10	333.0
221.00	13704.0
223.00	4561.0
224.00	44832.0
225.00	12402.0
226.10	1142.0
227.00	18048.0
227.90	2159.0
229.00	3548.0
230.00	584.0
231.00	1434.0
234.00	983.0
234.90	1418.0
235.90	970.0
237.00	1742.0
239.00	757.0
240.00	593.0
241.00	895.0
242.00	2330.0
243.00	2630.0
244.10	38304.0
245.00	4902.0
246.00	6820.0
246.90	1374.0
248.00	321.0
248.90	1266.0
251.00	363.0
251.80	310.0
253.00	768.0
255.00	191872.0
256.00	26624.0
257.00	2008.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

258.00	9415.0
259.00	1718.0
260.00	402.0
263.90	310.0
264.90	4186.0
266.00	560.0
270.00	302.0
270.90	350.0
272.00	503.0
273.00	5542.0
274.00	15242.0
275.00	79352.0
276.00	10914.0
277.00	6251.0
278.00	1521.0
281.00	370.0
283.00	721.0
284.00	568.0
285.00	796.0
292.00	323.0
293.00	1696.0
293.90	441.0
295.00	465.0
296.00	21936.0
297.00	2880.0
302.00	451.0
303.00	2777.0
303.90	588.0
310.00	313.0
314.00	1177.0
315.00	2373.0
316.00	1568.0
317.10	367.0
321.00	926.0
322.10	355.0
323.10	7656.0
324.00	1746.0
327.00	1307.0
328.00	703.0
332.00	535.0
332.90	901.0
334.00	4808.0
335.00	1194.0
341.00	1030.0
346.00	1598.0
347.00	325.0
352.00	2430.0
353.00	1882.0
354.00	2539.0
355.00	472.0
365.00	9573.0
366.00	1389.0
370.90	633.0
372.10	4147.0
373.00	992.0
382.90	1269.0
390.00	516.0
391.00	425.0
401.90	1545.0
403.00	2347.0
404.10	1073.0
420.90	2040.0
422.00	2152.0
423.00	16314.0
424.00	4159.0
424.90	460.0
441.00	46648.0
442.00	338304.0
443.00	65592.0
444.00	6401.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

m/z	Abundance
38.00	671.0
39.00	4166.0
40.00	903.0
44.00	4363.0
48.90	971.0
50.00	16079.0
51.00	63688.0
52.00	2983.0
54.90	380.0
56.00	2281.0
57.00	4553.0
58.00	364.0
61.00	777.0
62.00	937.0
63.00	2646.0
63.90	327.0
65.00	1422.0
68.00	1190.0
69.00	67176.0
70.00	490.0
73.00	617.0
74.00	6479.0
75.00	10614.0
76.00	3728.0
77.00	81432.0
78.00	5408.0
79.00	5126.0
80.00	3670.0
81.00	5030.0
82.00	1325.0
82.90	1365.0
84.00	518.0
84.90	1056.0
85.90	1986.0
87.00	845.0
88.00	459.0
91.00	1180.0
91.90	1470.0
93.00	8447.0
93.90	620.0
96.00	632.0
98.00	6892.0
98.90	5609.0
99.90	563.0
100.90	3387.0
103.00	1381.0
104.00	2128.0
104.90	1846.0
106.10	672.0
107.00	25576.0
108.00	3817.0
109.00	734.0
110.00	47480.0
111.00	7043.0
112.00	684.0
116.00	1260.0
117.00	19464.0
118.00	1439.0
121.90	1785.0
122.90	2130.0
123.90	1078.0
124.90	1085.0
126.10	555.0
127.00	92040.0
128.00	6497.0
129.00	33688.0
129.90	3177.0
130.90	620.0
132.90	477.0
133.90	1015.0
135.00	2904.0
136.00	1041.0
137.00	1663.0
137.80	316.0
140.00	516.0
141.00	4379.0
141.90	1481.0
143.00	1188.0
143.90	367.0
145.90	785.0
147.00	2374.0
148.00	5379.0
149.00	1083.0
150.20	333.0
151.10	573.0
151.70	359.0
153.00	1742.0
154.00	1287.0
155.00	2824.0
156.00	3760.0
157.00	936.0
157.90	907.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

158.90	669.0
159.90	1548.0
161.00	2247.0
162.00	584.0
165.00	1660.0
166.00	1422.0
167.00	9558.0
168.00	5680.0
169.00	846.0
171.00	434.0
172.00	811.0
173.00	1274.0
174.00	2434.0
175.00	4206.0
176.00	925.0
176.90	1944.0
177.90	646.0
179.00	7486.0
180.00	4879.0
181.00	2286.0
182.00	452.0
184.00	667.0
184.90	3468.0
186.00	28552.0
187.00	8097.0
187.90	1070.0
189.00	1645.0
189.90	309.0
191.00	1027.0
192.00	2742.0
193.00	2439.0
194.10	536.0
194.90	415.0
196.00	5988.0
197.00	1408.0
198.00	211328.0
199.00	14362.0
199.90	1161.0
201.50	1017.0
203.00	1284.0
204.00	6988.0
205.00	12914.0
206.10	56440.0
207.00	8075.0
208.00	2517.0
209.00	841.0
210.00	801.0
210.90	2261.0
211.90	329.0
215.00	635.0
216.00	1148.0
216.90	13848.0
217.90	1758.0
221.00	8583.0
221.60	2200.0
223.00	2936.0
224.00	29616.0
225.00	7529.0
226.10	726.0
227.00	12407.0
228.00	1399.0
228.90	2841.0
229.90	340.0
230.90	860.0
233.90	665.0
235.00	950.0
236.00	657.0
236.90	1156.0
239.00	553.0
239.90	441.0
241.00	668.0
242.00	1442.0
243.00	1870.0
244.00	25304.0
245.00	3543.0
245.90	4536.0
247.00	929.0
249.00	906.0
252.90	652.0
254.00	1286.0
255.00	131904.0
256.00	18048.0
257.00	1360.0
258.00	6977.0
259.00	1068.0
265.00	2429.0
266.00	449.0
271.00	344.0
272.10	317.0
273.00	3955.0
274.00	10175.0
275.00	59408.0
276.00	8039.0
277.00	4670.0
278.00	868.0
281.00	314.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

283.00	477.0
283.90	334.0
284.90	988.0
292.90	1125.0
293.90	309.0
295.00	399.0
296.00	15950.0
297.00	2491.0
303.00	1858.0
304.00	514.0
308.00	330.0
309.90	331.0
314.00	980.0
315.00	2065.0
316.00	1271.0
320.90	464.0
323.00	6090.0
324.00	1002.0
326.90	1213.0
327.90	521.0
331.90	359.0
332.90	806.0
334.00	3499.0
335.00	1022.0
341.00	601.0
346.00	1245.0
352.00	1871.0
353.00	1113.0
354.00	1612.0
355.10	437.0
365.00	7707.0
366.00	1083.0
370.10	307.0
370.90	397.0
372.00	3764.0
373.00	779.0
382.90	928.0
389.90	396.0
391.90	389.0
401.90	1099.0
403.00	2047.0
403.90	662.0
421.00	1722.0
422.10	1879.0
423.00	12763.0
424.00	3022.0
441.10	40720.0
442.00	269568.0
443.00	51888.0
444.00	3605.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253