

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062822\
 Data File : BN020563.D
 Acq On : 29 Jun 2022 10:35
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 29 13:57:48 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	1,4-Dioxane	0.524	0.475	9.4	117	0.00
3	n-Nitrosodimethylamine	0.478	0.389	18.6	108	0.00
4 S	2-Fluorophenol	1.110	0.938	15.5	107	0.00
5 S	Phenol-d6	1.327	1.153	13.1	113	0.00
6	bis(2-Chloroethyl)ether	1.142	1.015	11.1	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
8 S	Nitrobenzene-d5	0.324	0.278	14.2	116	-0.01
9	Naphthalene	1.195	1.148	3.9	123	-0.01
10	Hexachlorobutadiene	0.218	0.204	6.4	119	-0.01
11 SURR	2-Methylnaphthalene-d10	0.686	0.634	7.6	121	0.00
12	2-Methylnaphthalene	0.794	0.749	5.7	121	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
14 S	2,4,6-Tribromophenol	0.151	0.121	19.9	113	0.00
15 S	2-Fluorobiphenyl	1.651	1.607	2.7	120	-0.01
16	Acenaphthylene	1.923	1.770	8.0	122	-0.01
17	Acenaphthene	1.308	1.259	3.7	121	-0.01
18	Fluorene	1.707	1.606	5.9	120	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	-0.01
20	4,6-Dinitro-2-methylphenol	0.051	0.032	37.3#	102	0.01
21	4-Bromophenyl-phenylether	0.233	0.216	7.3	121	-0.01
22	Hexachlorobenzene	0.257	0.248	3.5	126	0.00
23	Atrazine	0.183	0.144	21.3	106	0.00
24	Pentachlorophenol	0.061	0.053	13.1	143	0.00
25	Phenanthrene	1.349	1.258	6.7	121	-0.01
26	Anthracene	1.163	1.030	11.4	119	0.00
27 SURR	Fluoranthene-d10	1.188	1.088	8.4	121	0.00
28	Fluoranthene	1.471	1.374	6.6	123	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
30	Pyrene	1.694	1.718	-1.4	120	0.00
31 S	Terphenyl-d14	0.956	0.952	0.4	117	0.00
32	Benzo(a)anthracene	1.429	1.278	10.6	109	0.00
33	Chrysene	1.582	1.552	1.9	117	0.00
34	Bis(2-ethylhexyl)phthalate	0.805	0.627	22.1	101	0.00
35 I	Perylene-d12	1.000	1.000	0.0	95	-0.02
36	Indeno(1,2,3-cd)pyrene	1.782	1.523	14.5	85	-0.02
37	Benzo(b)fluoranthene	1.541	1.566	-1.6	101	-0.01
38	Benzo(k)fluoranthene	1.641	1.663	-1.3	103	-0.01
39 C	Benzo(a)pyrene	1.337	1.275	4.6	95	-0.01
40	Dibenzo(a,h)anthracene	1.429	1.210	15.3	85	-0.02
41	Benzo(g,h,i)perylene	1.561	1.343	14.0	84	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4