

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062822\
 Data File : BN020562.D
 Acq On : 29 Jun 2022 09:59
 Operator : CG/JU
 Sample : N3457-08
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D5-20220620

Quant Time: Jun 29 13:57:31 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 02:03:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	2999	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	9426	0.400	ng	# 0.00
13) Acenaphthene-d10	14.367	164	5737	0.400	ng	0.00
19) Phenanthrene-d10	17.113	188	12183	0.400	ng	0.00
29) Chrysene-d12	21.306	240	11054	0.400	ng	# 0.00
35) Perylene-d12	23.583	264	8364	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	788	0.095	ng	0.00
5) Phenol-d6	6.966	99	455	0.046	ng	0.02
8) Nitrobenzene-d5	8.907	82	1511	0.198	ng	0.01
11) 2-Methylnaphthalene-d10	12.117	152	4397	0.272	ng	0.00
14) 2,4,6-Tribromophenol	15.861	330	462	0.214	ng	0.00
15) 2-Fluorobiphenyl	12.998	172	6336	0.268	ng	0.00
27) Fluoranthene-d10	19.143	212	12091	0.334	ng	0.00
31) Terphenyl-d14	19.742	244	8509	0.322	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.225	88	17790	4.526	ng	Qvalue 93
6) bis(2-Chloroethyl)ether	7.183	93	259	0.030	ng	97
34) Bis(2-ethylhexyl)phtha...	21.225	149	926	0.042	ng	# 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062822\
Data File : BN020562.D
Acq On : 29 Jun 2022 09:59
Operator : CG/JU
Sample : N3457-08
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D5-20220620


Quant Time: Jun 29 13:57:31 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 02:03:21 2022
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.739 min Scan# 64
 Delta R.T. -0.000 min
 Lab File: BN020562.D
 Acq: 29 Jun 2022 09:59


Instrument :
 BNA_N
 ClientSampleId :
 RE132D5-20220620





Tgt Ion:152 Resp: 2999
 Ion Ratio Lower Upper
 152 100
 150 155.6 127.9 191.9
 115 65.2 51.4 77.2




#2
 1,4-Dioxane
 Concen: 4.526 ng
 RT: 3.225 min Scan# 20
 Delta R.T. -0.007 min
 Lab File: BN020562.D
 Acq: 29 Jun 2022 09:59



Tgt Ion: 88 Resp: 17790
 Ion Ratio Lower Upper
 88 100
 43 33.1 30.2 45.2
 58 84.6 73.0 109.6









#4
2-Fluorophenol
Concen: 0.095 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59




Instrument :
BNA_N
ClientSampleId :
RE132D5-20220620


Tgt Ion:112 Resp: 788
Ion Ratio Lower Upper
112 100
64 64.1 49.1 73.7
63 35.3 26.3 39.5



#5
Phenol-d6
Concen: 0.046 ng
RT: 6.966 min Scan# 538
Delta R.T. 0.022 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59

Tgt Ion: 99 Resp: 455
Ion Ratio Lower Upper
99 100
42 12.3 19.3 28.9#
71 31.6 27.5 41.3









#6
bis(2-Chloroethyl)ether
Concen: 0.030 ng
RT: 7.183 min Scan# 566
Delta R.T. 0.014 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59




Instrument :
BNA_N
ClientSampleId :
RE132D5-20220620


Tgt Ion: 93 Resp: 259
Ion Ratio Lower Upper
93 100
63 85.3 67.4 101.0
95 28.6 26.5 39.7



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.519 min Scan# 931
Delta R.T. -0.000 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59

Tgt Ion: 136 Resp: 9426
Ion Ratio Lower Upper
136 100
137 11.9 9.0 13.6
54 12.0 6.2 9.2#
68 8.0 4.4 6.6#









#8
 Nitrobenzene-d5
 Concen: 0.198 ng
 RT: 8.907 min Scan# 71
 Delta R.T. 0.011 min
 Lab File: BN020562.D
 Acq: 29 Jun 2022 09:59




Instrument :
 BNA_N
 ClientSampleId :
 RE132D5-20220620


Tgt Ion: 82 Resp: 1511
 Ion Ratio Lower Upper
 82 100
 128 44.8 35.0 52.6
 54 65.5 46.2 69.4



#11
 2-Methylnaphthalene-d10
 Concen: 0.272 ng
 RT: 12.117 min Scan# 1184
 Delta R.T. -0.004 min
 Lab File: BN020562.D
 Acq: 29 Jun 2022 09:59

Tgt Ion: 152 Resp: 4397
 Ion Ratio Lower Upper
 152 100
 151 21.7 16.2 24.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.367 min Scan# 1461
Delta R.T. -0.000 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59




Instrument :
BNA_N
ClientSampleId :
RE132D5-20220620


Tgt Ion:164 Resp: 5737
Ion Ratio Lower Upper
164 100
162 103.9 82.1 123.1
160 47.9 38.6 58.0



#14
2,4,6-Tribromophenol
Concen: 0.214 ng
RT: 15.861 min Scan# 1601
Delta R.T. -0.000 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59

Tgt Ion:330 Resp: 462
Ion Ratio Lower Upper
330 100
332 198.3 77.6 116.4#
141 58.7 41.0 61.4









#15
2-Fluorobiphenyl
Concen: 0.268 ng
RT: 12.998 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59




Instrument :
BNA_N
ClientSampleId :
RE132D5-20220620


Tgt Ion:172 Resp: 6336
Ion Ratio Lower Upper
172 100
171 34.9 27.2 40.8
170 23.2 18.2 27.4



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.113 min Scan# 1723
Delta R.T. -0.000 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59

Tgt Ion:188 Resp: 12183
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.0 7.4 11.2









#27
Fluoranthene-d10
Concen: 0.334 ng
RT: 19.143 min Scan# 2080
Delta R.T. -0.004 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59




Instrument :
BNA_N
ClientSampleId :
RE132D5-20220620


Tgt Ion:212 Resp: 12091
Ion Ratio Lower Upper
212 100
106 15.8 12.6 19.0
104 8.8 7.2 10.8



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.306 min Scan# 2430
Delta R.T. -0.000 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59

Tgt Ion:240 Resp: 11054
Ion Ratio Lower Upper
240 100
120 12.4 14.6 21.8#
236 27.1 22.6 33.8









#31
 Terphenyl-d14
 Concen: 0.322 ng
 RT: 19.742 min Scan# 21
 Delta R.T. -0.013 min
 Lab File: BN020562.D
 Acq: 29 Jun 2022 09:59




Instrument :
 BNA_N
 ClientSampleId :
 RE132D5-20220620


Tgt Ion:244 Resp: 8509
 Ion Ratio Lower Upper
 244 100
 212 9.2 7.1 10.7
 122 0.0 0.0 0.0



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.042 ng
 RT: 21.225 min Scan# 2421
 Delta R.T. -0.009 min
 Lab File: BN020562.D
 Acq: 29 Jun 2022 09:59


Tgt Ion:149 Resp: 926
 Ion Ratio Lower Upper
 149 100
 167 26.2 21.7 32.5
 279 4.3 2.2 3.2#





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.583 min Scan# 2961
Delta R.T. -0.018 min
Lab File: BN020562.D
Acq: 29 Jun 2022 09:59

Instrument :
BNA_N
ClientSampleId :
RE132D5-20220620



Tgt Ion:264 Resp: 8364
Ion Ratio Lower Upper
264 100
260 26.2 20.9 31.3
265 36.6 28.0 42.0

