

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062921\
 Data File : BN015207.D
 Acq On : 30 Jun 2021 04:10
 Operator : CG/JU
 Sample : M2783-09
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE115D2-20210617

Manual Integrations
 APPROVED

mohammad
 6/30/2021 2:15:14 PM

Quant Time: Jun 30 06:20:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 29 19:13:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	16310	0.400	ng	# 0.00
7) Naphthalene-d8	10.493	136	73676	0.400	ng	#-0.01
13) Acenaphthene-d10	14.346	164	34000	0.400	ng	0.00
19) Phenanthrene-d10	17.091	188	67257	0.400	ng	0.00
29) Chrysene-d12	21.292	240	62607	0.400	ng	#-0.01
35) Perylene-d12	23.537	264	52228	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.371	112	4491	0.127	ng	0.03
5) Phenol-d6	6.911	99	4191	0.078	ng	0.00
8) Nitrobenzene-d5	8.870	82	26285	0.390	ng	0.00
11) 2-Methylnaphthalene-d10	12.088	152	42226	0.349	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	2657	0.360	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	63934	0.448	ng	-0.01
27) Fluoranthene-d10	19.128	212	78589	0.384	ng	0.00
31) Terphenyl-d14	19.734	244	75914	0.549	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.352	88	23203m	4.901	ng	Qvalue
6) bis(2-Chloroethyl)ether	7.169	93	2598	0.050	ng	# 83
34) Bis(2-ethylhexyl)phtha...	21.218	149	4866	0.301	ng	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062921\
Data File : BN015207.D
Acq On : 30 Jun 2021 04:10
Operator : CG/JU
Sample : M2783-09
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE115D2-20210617

Manual Integrations
APPROVED

mohammad
6/30/2021 2:15:14 PM

Quant Time: Jun 30 06:20:34 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 29 19:13:09 2021
Response via : Initial Calibration

