

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062921\
 Data File : BN015212.D
 Acq On : 30 Jun 2021 07:10
 Operator : CG/JU
 Sample : M2783-14
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP03-20210617

Manual Integrations
 APPROVED

mohammad
 6/30/2021 2:15:26 PM

Quant Time: Jun 30 08:37:49 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 29 19:13:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	16854	0.400	ng	# 0.00
7) Naphthalene-d8	10.493	136	77411	0.400	ng	#-0.01
13) Acenaphthene-d10	14.346	164	35526	0.400	ng	0.00
19) Phenanthrene-d10	17.091	188	72031	0.400	ng	0.00
29) Chrysene-d12	21.292	240	66430	0.400	ng	#-0.01
35) Perylene-d12	23.540	264	53229	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.371	112	3894	0.107	ng	0.03
5) Phenol-d6	6.911	99	3655	0.066	ng	0.00
8) Nitrobenzene-d5	8.870	82	25353	0.358	ng	0.00
11) 2-Methylnaphthalene-d10	12.088	152	40267	0.317	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	2372	0.336	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	60360	0.405	ng	-0.01
27) Fluoranthene-d10	19.128	212	78881	0.360	ng	0.00
31) Terphenyl-d14	19.734	244	69891	0.476	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.352	88	6001m	1.227	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.218	149	5671	0.305	ng	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN062921\
Data File : BN015212.D
Acq On : 30 Jun 2021 07:10
Operator : CG/JU
Sample : M2783-14
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP03-20210617

Manual Integrations
APPROVED

mohammad
6/30/2021 2:15:26 PM

Quant Time: Jun 30 08:37:49 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 29 19:13:09 2021
Response via : Initial Calibration

